

ABN 43 064 478 842

231 osborne avenue clayton south, vic 3169
PO box 1548, clayton south, vic 3169
t 03 9265 7400 f 03 9558 0875
freecall 1800 680 680
www.tmgtestequipment.com.au

Test & Measurement

Complimentary Reference Material

sales
rentals
calibration
repair
disposal
This PDF has been made available as a complimentary service for you to assist in evaluating this model for your testing requirements.
TMG offers a wide range of test equipment solutions, from renting short to long term, buying refurbished and purchasing new. Financing options, such as Financial Rental, and Leasing are also available on application.
TMG will assist if you are unsure whether this model will suit your requirements.
Call TMG if you need to organise repair and/or calibrate your unit.
If you click on the "Click-to-Call" logo below, you can all us for FREE!



Disclaimer:

All trademarks appearing within this PDF are trademarks of their respective owners.



Tektronix

THS 710 & THS 720 TekScope[™] User Manual

070-9247-02

This document applies to firmware version 1.00 and above.

Copyright © Tektronix, Inc. 1995. All rights reserved.

Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supercedes that in all previously published material. Specifications and price change privileges reserved.

Printed in the U.S.A.

Tektronix, Inc., P.O. Box 1000, Wilsonville, OR 97070-1000

TEKTRONIX and TEK are registered trademarks of Tektronix, Inc.

Tek Secure is a registered trademark of Tektronix, Inc.

TekTools, TekScope, and IsolatedChannel are trademarks of Tektronix, Inc.

WARRANTY

Tektronix warrants that this product will be free from defects in materials and workmanship for a period of three (3) years from the date of shipment. If any such product proves defective during this warranty period, Tektronix, at its option, either will repair the defective product without charge for parts and labor, or will provide a replacement in exchange for the defective product.

In order to obtain service under this warranty, Customer must notify Tektronix of the defect before the expiration of the warranty period and make suitable arrangements for the performance of service. Customer shall be responsible for packaging and shipping the defective product to the service center designated by Tektronix, with shipping charges prepaid. Tektronix shall pay for the return of the product to Customer if the shipment is to a location within the country in which the Tektronix service center is located. Customer shall be responsible for paying all shipping charges, duties, taxes, and any other charges for products returned to any other locations.

This warranty shall not apply to any defect, failure or damage caused by improper use or improper or inadequate maintenance and care. Tektronix shall not be obligated to furnish service under this warranty a) to repair damage resulting from attempts by personnel other than Tektronix representatives to install, repair or service the product; b) to repair damage resulting from improper use or connection to incompatible equipment; or c) to service a product that has been modified or integrated with other products when the effect of such modification or integration increases the time or difficulty of servicing the product.

THIS WARRANTY IS GIVEN BY TEKTRONIX WITH RESPECT TO THIS PRODUCT IN LIEU OF ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED. TEKTRONIX AND ITS VENDORS DISCLAIM ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. TEKTRONIX' RESPONSIBILITY TO REPAIR OR REPLACE DEFECTIVE PRODUCTS IS THE SOLE AND EXCLUSIVE REMEDY PROVIDED TO THE CUSTOMER FOR BREACH OF THIS WARRANTY. TEKTRONIX AND ITS VENDORS WILL NOT BE LIABLE FOR ANY INDIRECT, SPECIAL, INCIDENTAL, OR CONSEQUENTIAL DAMAGES IRRESPECTIVE OF WHETHER TEKTRONIX OR THE VENDOR HAS ADVANCE NOTICE OF THE POSSIBILITY OF SUCH DAMAGES.

Service Assurance

If you have not already purchased Service Assurance for this product, you may do so at any time during the product's warranty period. Service Assurance provides Repair Protection and Calibration Services to meet your needs.

Repair Protection extends priority repair services beyond the product's warranty period; you may purchase up to three years of Repair Protection.

Calibration Services provide annual calibration of your product, standards compliance and required audit documentation, recall assurance, and reminder notification of scheduled calibration. Coverage begins upon registration; you may purchase up to five years of Calibration Service.

Service Assurance Advantages

- Priced well below the cost of a single repair or calibration
- Avoid delays for service by eliminating the need for separate purchase authorizations from your company
- Eliminates unexpected service expenses

For Information and Ordering

For more information or to order Service Assurance, contact your Tektronix representative and provide the information below. Service Assurance may not be available in locations outside the United States of America.

Name	VISA or Master Card expiration date or
Company	purchase order number
Address	Repair Protection (1,2 or 3 years)
City, State, Postal code	Calibration Service (1,2,3,4 or 5 years)
Country	Instrument model and serial number
Phone	Instrument purchase date

Table of Contents

General Safety Summary	iii
Preface	v
In This Manual	v
Conventions	vi

Getting Started

Product Description	1–1
Replacing the Battery Pack	1–5
Using External Power	1–7
Using the Tilt Stand	1–8
Functional Check	1–9

Operating Basics

-1
-1
-6
-7
-8
0
13
4
6
8
20
22
24
26
28
30
32
34
36
38

Reference

Introduction to Reference	3–1
ACQUIRE	3–3
AUTORANGE	3–8
CURSOR	3–11
DISPLAY	3–13
HARD COPY	3–17
HOLD	3–19
HORIZONTAL Controls	3–20
MEAS	3–24
METER Mode	3–31
SAVE/RECALL	3–37
SCOPE Mode	3–40
TRIGGER Controls	3–46
UTILITY	3–53
VERTICAL Controls	3–59

Appendices

Appendix A: Specifications	A-1
Appendix B: Factory Setup	B-1
Appendix C: Accessories	C-1
Appendix D: Performance VerificationTest RecordPerformance Verification Procedures	D-1 D-2 D-4
Appendix E: General Care and CleaningGeneral CareCleaning	E–1 E–1 E–1

Glossary and Index

Glossary Index

General Safety Summary

Review the following safety precautions to avoid injury and prevent damage to this product or any products connected to it.

Only qualified personnel should perform service procedures.

Injury Precautions

Use Proper Power Cord. To avoid fire hazard, use only the power cord specified for this product.

Avoid Electric Overload. To avoid injury or fire hazard, do not apply potential to any input, including the common inputs, that varies from ground by more than the maximum rating for that input.

Avoid Electric Shock. To avoid injury or loss of life, do not connect or disconnect probes or test leads while they are connected to a voltage source.

Do Not Operate Without Covers. To avoid electric shock or fire hazard, do not operate this product with covers or panels removed.

Do Not Operate in Wet/Damp Conditions. To avoid electric shock, do not operate this product in wet or damp conditions.

Do Not Operate in an Explosive Atmosphere. To avoid injury or fire hazard, do not operate this product in an explosive atmosphere.

Product Damage Precautions

Use Proper Power Source. Do not operate this product from a power source that applies more than the voltage specified.

Do Not Operate With Suspected Failures. If you suspect there is damage to this product, have it inspected by qualified service personnel.

Safety Terms and Symbols

Terms in This Manual. These terms may appear in this manual:



WARNING. Warning statements identify conditions or practices that could result in injury or loss of life.



CAUTION. Caution statements identify conditions or practices that could result in damage to this product or other property.

Terms on the Product. These terms may appear on the product:

DANGER indicates an injury hazard immediately accessible as you read the marking.

WARNING indicates an injury hazard not immediately accessible as you read the marking.

CAUTION indicates a hazard to property including the product.

Symbols on the Product. These symbols may appear on the product:





DANGER High Voltage

Protective Ground (Earth) Terminal

ATTENTION

Refer to Manual



Double Insulated

Certifications and Compliances

CSA Certified AC Adapter. CSA Certification includes the AC adapters appropriate for use in the North America power network. All other AC adapters supplied are approved for the country of use.

Compliances. Consult the product specifications for Overvoltage Category and Safety Class.

Preface

This User Manual describes the capabilities, operation, and applications of the THS 710 and THS 720 TekScope instruments.

In This Manual

The following table shows you where to find information in this manual.

If you are looking for:	Turn to:
Product overview	Product Description on page 1–1
Details about a product feature	Reference on page 3–1
	Look up the front-panel button for the feature
Application examples	Application Examples on page 2–13
Operation instructions	<i>Understanding the Front Panel</i> on page 2–1
Information about battery operation	<i>Changing the Battery Pack</i> on page 1–5
Information about using external power	Using External Power on page 1–7
Information about making a hard copy	HARD COPY on page 3–17
Technical specifications	Specifications appendix on page A-1
Recommended accessories	Accessories appendix on page C–1

Conventions

TekScope instrument setups are shown in tables. The *Application Examples* and *Performance Verification* sections use tables to show specific setups. The *Reference* section uses similar tables to show the complete contents of the menu system.

The header of each table contains icons that represent the controls and menu items used to set up the instrument. To make a specific setup, read the table from left to right and then from top to bottom as shown below. The table contains the symbol "—" if no action is required.



Getting Started

Getting Started

In addition to a brief product description, this chapter covers the following topics:

- How to change the battery pack
- How to use external power
- How to use the tilt stand
- How to perform a quick functional check

Product Description

The THS 710 and THS 720 TekScope instruments combine a two-channel oscilloscope and a digital multimeter (DMM) in a rugged, handheld package.

General Features

- Battery power or external power
- High-resolution, high-contrast display with temperature compensation for clear visibility over a wide temperature range
- Onboard waveform, data, and setup storage
- RS-232 communication port to load setups, download waveforms, and make hard copies
- Fully programmable through the RS-232 communication port



Oscilloscope Features

The TekScope instrument is a powerful, two-channel oscilloscope with the following features:

- Autoranging for quick setup and hands-free operation
- 100 MHz (THS 720) or 60 MHz (THS 710) bandwidth with selectable 20 MHz bandwidth limit
- 500 MS/s (THS 720) or 250 MS/s (THS 710) sample rate and 2,500 point record length
- Separate digitizers for each channel (both channels always acquire simultaneously)
- Waveform averaging and enveloping with hardware peak detection
- Digital Real Time digitizing (up to five-times oversampling), sin(x)/x interpolation, and peak-detect acquisition to limit the possibility of aliasing
- Independently isolated channels to allow improved safety for measurements to 1000 V_{RMS} while floating up to 600 V_{RMS} using P5102 probes
- Cursors and 21 continuously updated, automatic measurements
- Simultaneous oscilloscope and meter operation on the same or on separate signals
- Advanced pulse and video trigger capability



Meter Features

The TekScope instrument is also a full-featured DMM with the following features:

- True RMS VAC, VDC, Ω , continuity, and diode-check functions
- Autoranging or manual ranging
- Data logger plot of meter measurements over a period of time
- Max, min, delta max-min, relative-delta, and average statistics in the readout
- Bar graph for an "analog meter" feel
- Capability to float independently to 600 V_{RMS}
- Overvoltage indicator warns when an overvoltage is applied to the input

Input and Output Connectors

All input and output connectors are located on the top and side panels as shown below.

NOTE. See the back of the instrument for maximum voltage ratings.



Top panel



Side panel

PROBE COMP \approx 5V Π

RS-232

Replacing the Battery Pack

For portable operation, use the rechargeable battery pack.

You can replace the battery pack without losing any saved information. The current setup, saved setups, saved waveforms, and saved data are all stored in nonvolatile memory that does not depend on battery power.



CAUTION. To prevent loss of saved information, set the ON/STBY switch to STBY before removing the battery back.



Battery Life

From a full charge, you can operate the TekScope instrument continuously for approximately two hours. You can extend the battery life by using automatic Power Off Time-out or Backlight Time-out. Refer to page 3–56 for a description of these features.

The TekScope instrument turns off automatically when the battery runs low. A low-battery message appears in the display about ten minutes before the automatic shutdown.

Nickel-cadmium batteries can lose capacity permanently if not allowed to discharge completely. Whenever possible, allow the battery to discharge completely before you recharge it to minimize this capacity loss.

Charging the Battery Pack

Before using the battery for the first time, it must be charged. You can use external power to charge the battery pack while it is in the TekScope instrument. Or you can charge the battery pack with the optional external battery charger. Typical battery charging times are listed below.

Configuration	Typical Charging Time
Battery pack in TekScope instrument	20 hours
Battery pack in external charger	1.5 hours



CAUTION. To avoid loss of saved information when a battery pack is not installed, set the ON/STBY control to STBY before disconnecting external power.

Using External Power

Using external power from the AC adapter or cigarette lighter adapter has the following advantages:

- Saves internal battery power for portable operation later
- Charges internal battery pack
- Allows extended operation; the Standby Time-out and Backlight Time-out features are automatically disabled when external power is used
- Maintains floating measurement capability of the oscilloscope channels and DMM

Attach the external power source as shown below.

The DC INPUT disconnects itself if an overvoltage is applied. If this occurs, disconnect and then reconnect the AC adapter or cigarette lighter adapter to resume operation from external power.





CAUTION. To avoid overheating, do not connect external power while the instrument is in a confined space, such as in the soft case.

Using the Tilt Stand

A built-in tilt stand folds out and snaps back into place when not in use. For benchtop use, lock the tilt stand in place with the hinged flap. To hang the TekScope instrument over a nail, rotate the tilt stand 180°. You can also extend the hinged flap as shown to hang the instrument from a ladder rung or over the top of a door.



Functional Check

After you install batteries or connect external power, you can perform this quick functional check to verify that your TekScope instrument is operating correctly.

- 1. Press the **ON/OFF** button to turn on the TekScope instrument.
- 2. After a few seconds, you should see a window with the message Power-On self check PASSED. Press the **CLEAR MENU** button.
- 3. Press the SCOPE button.
- **4.** Connect the oscilloscope probe to the channel 1 input BNC. Attach the probe tip and reference lead to the PROBE COMP connectors on the right side of the TekScope instrument.



5. Press the **AUTORANGE** button. Within a few seconds, you should see a square wave in the display (approximately 1.2 kHz).

If you want, repeat steps 4 and 5 for channel 2 of the oscillo-scope.

- 6. Press the **METER** button.
- 7. Press the VDC bezel button
- 8. Press the AUTORANGE button.
- **9.** Connect meter leads to the TekScope instrument and touch the meter lead tips to the PROBE COMP output as shown below.



10. Verify that the TekScope instrument measures an average DC voltage of 2.5 ± 0.25 V.

Operating Basics

Functional Overview

This section covers the following topics:

- Understanding the front panel
- Using scope mode
- Using meter mode
- Connecting and using the probes
- Taking floating measurements

You can find specific information about each of the controls in the *Reference* chapter of this manual.

Understanding the Front Panel

The front panel has buttons for the functions you use most often and menus to access more specialized functions. With the autorange feature, you can setup the TekScope instrument automatically in both scope and meter modes.

Using the Menu System

To use the menu system, follow the steps shown on the next two pages.



1. Press a front-panel button to display the menu you want to use.

2. Press a bezel button to choose a menu item. If a pop-up menu appears, continue to press the bezel button to choose an item in the pop-up menu. You may need to press the Select Page bezel button to access additional menu items.



3. Certain menu choices require you to set a numerical parameter to complete the setup. Use the ⁺/– rocker to adjust the parameter value or press the TOGGLE button to reset the parameter to its default value.



4. If the OK bezel button is displayed, press it to confirm your choice.



Using the Dedicated Buttons

You can use the dedicated buttons below to take direct actions. These buttons do not require the use of menus.



- 1. HARD COPY. Initiates a hard copy using the RS-232 port.
- **2.** HOLD. Stops/restarts oscilloscope acquisition or holds/resets meter readout.
- **3**. AUTORANGE. Selects oscilloscope or meter Autorange function.

- 4. CLEAR MENU. Clears menu from display.
- 5. TRIGGER LEVEL. Adjusts trigger level.
- **6**. SET LEVEL TO 50%. Sets trigger level to midpoint of oscilloscope waveform.
- 7. HORIZONTAL POSITION. Adjusts oscilloscope waveform horizontal position.
- 8. MAG. Turns 10X horizontal magnification on and off.
- **9**. SEC/DIV. Adjusts horizontal scale factor for oscilloscope or data logger.
- **10**. VERTICAL POSITION. Adjusts oscilloscope waveform vertical position.
- **11.** WAVEFORM OFF. Removes selected oscilloscope waveform from display.
- **12.** VOLTS/DIV. Adjusts oscilloscope vertical scale factor or meter range.
- **13.** ON/STBY. Selects on or standby. Does not disconnect power from the instrument.
- **14**. CH 1, CH 2, MATH, REF A, REF B. Displays waveform and chooses selected waveform.
- **15.** SCOPE. Selects scope mode.
- **16.** METER. Selects meter mode.

Using Scope Mode

Press the front-panel SCOPE button to enter scope mode. Then, press AUTORANGE to set the vertical, horizontal, and trigger automatically for a usable display.

The scope-mode display, shown below, is divided into four sections. Refer to *SCOPE Mode* on page 3–40 for a description of each section.



Using Meter Mode



Press the front-panel METER button to enter meter mode. Press one of the bezel buttons to choose a meter function and then press AUTORANGE to set the range automatically.

The meter-mode display, shown below, is divided into three sections. Refer to *METER Mode* on page 3–31 for a description of each section and more information about the data logger and bar graph.



Graticule area

Compensating the Oscilloscope Probes



To maintain signal fidelity, you must compensate each voltage probe for the channel input it is connected to.

1. Connect the oscilloscope probe and then press AUTORANGE.



2. Check the shape of the displayed waveform.



3. If necessary, adjust the probe for correct compensation.



4. Repeat these steps for the other probe and channel.

Taking Floating Measurements

This section covers important issues to consider when taking floating measurements.

Architecture is Important

For taking floating measurements, the TekScope instrument has an architectural difference from most other oscilloscopes. The channel 1, channel 2, and DMM inputs are isolated from the main chassis and from each other. This architecture allows independent floating measurements with channel 1, channel 2, and the DMM.



Many handheld oscilloscope/DMM products have the architecture shown below, which shares a common reference for the oscilloscope channels and DMM. With this architecture, all input signals must have the same voltage reference when you take any multi-channel measurements.



Most bench-top oscilloscopes share the above architecture but without the insulated case. Without differential preamplifiers or external signal isolators, bench-top oscillscopes are not suitable for taking floating measurements.
Attach the Reference Leads Correctly

If you are using both of the oscilloscope channels, you must attach the probe reference lead for each channel directly to your circuit. These attachments are required because the oscilloscope channels are electrically isolated; they do not share a common chassis connection. Use the shortest possible reference lead with each probe to maintain good signal fidelity. If you are also using the DMM, you must also attach the DMM common lead to your circuit for the same reason as above.

The probe reference lead presents a higher capacitive load to the circuit-under-test than the probe tip. When taking a floating measurement between two nodes of a circuit, attach the probe reference lead to the lowest impedance or least dynamic of the two nodes.

Beware of High Voltages

Understand the voltage ratings for the probes you are using and do not exceed those ratings. Two ratings are important to know and understand:

- The maximum measurement voltage from the probe tip to the probe reference lead
- The maximum floating voltage from the probe reference lead to earth ground

These two voltage ratings depend on the probe and your application. Refer to *Specifications* beginning on page A–1 for more information.



WARNING. To prevent electrical shock, do not exceed the measurement or floating voltage ratings for the oscilloscope input BNC connector, probe tip, probe reference lead, DMM input connector, or DMM lead.

Application Examples

This section presents a series of application examples. These simplified examples highlight the features of the TekScope instrument and give you ideas about using it to solve your own test problems.

The first two examples demonstrate basic scope and meter operation. The remaining examples provide an overview of applications that cover the following areas:

- Digital circuit testing
- Analog circuit testing
- Power electronic testing
- Motor testing
- Power quality monitoring
- Video signal testing

Displaying an Unknown Signal

You need to see a signal in a circuit, but you have no previous knowledge of the signal amplitude, frequency, or shape. Connect the TekScope instrument to quickly display the signal.



Setup to Display an Unknown Signal



The autorange feature sets the vertical, horizontal, and trigger automatically for a usable display. If the signal changes, the setup tracks those changes.

Going Further

If the autorange setup does not display the waveform exactly the way you like, you can easily change the setup. Press any of the buttons below to cancel autorange mode and modify the setup:

- VOLTS/DIV
- SEC/DIV
- TRIGGER LEVEL
- SET TRIGGER LEVEL TO 50%

Measuring Resistance

You need to probe a circuit to measure point-to-point resistance. Connect the TekScope instrument to measure a variety of resistance values.



Setup to Measure Resistance



Going Further

If a noisy environment causes an unstable resistance measurement, use the Average statistic to average the measurements. Refer to page 3–29 for more information.

You can use the TekScope instrument as a continuity checker. With the setup below, it beeps when the measured resistance is 50 Ω or less (typical).

METER METER				TOGGLE +
METER	_	n)) (continuity)	_	_

You can also use the TekScope instrument as a semiconductor-junction checker. Use the setup below to measure the voltage drop across the junction. The open-circuit voltage is limited to about 4.8 V to prevent damage to reverse-biased junctions.

				TOGGLE +
METER	_	-}+ (diode)	_	_

Measuring the Frequency of a Clock Signal

You suspect that the frequency of a TTL clock signal is out of tolerance. Connect the TekScope instrument to the signal to display it and measure its frequency.



Setup to Measure Clock Frequency

METER () SCOPE				TOGGLE +
SCOPE	AUTO RANGE	_	_	_
	MEAS	Select Measmnt for Ch1	Frequency	
		OK Select Measrmnt	_	

Going Further

You can add peak-to-peak and duty cycle measurements to the display with the following additional steps:

METER METER				TOGGLE +
SCOPE	MEAS	Select Page	—	_
		Select Measrmnt for Ch1	Positive Duty Cycle	
		OK Select Measrmt	_	
		Select Page (press once)		
		Select Measrmnt for Ch1	Pk-Pk	
		OK Select Measrmt	_	

Measuring Propagation Delay

You suspect that the memory timing in a microprocessor circuit is marginal. Set up the TekScope instrument to measure the propagation delay between the chip-select signal and the data output of the memory device.



				TOGGLE +
SCOPE	CH 1	_	_	_
	CH 2			
	AUTO RANGE ¹			
	CURSOR	Function	V Bars	Adjust first cursor, press TOGGLE, and then adjust second cursor

Setup to Measure Propagation Delay

¹ If necessary, adjust the SEC/DIV rocker to optimize display for the propagation delay measurement.

Set one cursor to the active edge of the chip-select signal and the second cursor to the data output transition. Read the propagation delay in the cursor readout.

Going Further

The above example uses cursors to take relative timing measurements (Δ -seconds) between two different waveforms. If you are measuring just one waveform, select the **Paired** cursor function to measure Δ -volts and Δ -seconds at the same time.

Triggering on a Missing Data Pulse

A positive-going TTL data pulse, $20 \ \mu s$ wide, should occur at least once every millisecond. The circuit is not working correctly and you suspect an occasional missing pulse. Set up the TekScope instrument to find the missing pulse.



				-			
++++	++++		++++		 -++++	-++++	

METER (SCOPE)				TOGGLE +
SCOPE	AUTO RANGE	_	_	_
	TRIGGER	Туре	Pulse	
	MENU	Source	Ch1	
		Polarity and Width	Negative	Set width to 1 ms
		Trigger When	Greater Than Width	_
		Mode	Normal	

Setup to Find Missing Data Pulse

The TekScope instrument triggers if the signal remains in the low state longer than 1 ms. If it does, you have found an occurrence of a missing pulse.

Going Further

You can extend this application in several ways:

- If the pulses are supposed to be periodic (1 ms period) and you suspect an occasional extra pulse, change the width setting to **980 µs** and the **Trigger When** submenu to **Less Than Width**. With this setup, TekScope instrument triggers if the spacing between the pulses ever drops below 980 µs, which indicates an occurrence of an extra pulse.
- Use the second channel to find the cause of the problem. You can correlate the cause and the effect because the TekScope instrument always acquires both channels at exactly the same time.

Detecting Narrow Glitches

An elapsed-time counter circuit operates from a precision, 1 kHz square wave, clock signal supplied by another source. Occasionally, the counter counts too fast. You suspect glitches in the clock signal are causing the problem. Set up the TekScope instrument to look for glitches in the clock signal.



Setup to Detect Narrow Glitches



Monitor the clock signal for several minutes. In Envelope acquisition mode, the TekScope instrument displays the 1 kHz square wave clock signal plus intermittent glitches that are as narrow as 8 ns.

Going Further

You can trigger on the glitch itself with the following setup:

METER () SCOPE				TOGGLE +
SCOPE	TRIGGER MENU	Туре	Pulse	—
		Source	Ch1	
		Polarity and Width	Positive	Set width to 500 µs
		Trigger When	Less Than Width	_
		Mode	Normal	

A triggers occurs only if a positive pulse narrower than 500 μs (half the period of the clock signal) is detected.

Testing a Switching Transistor Drive Circuit

You need to evaluate the gate-drive circuit for a power FET (field-effect transistor) in a switching power supply. The gate-drive timing circuit is referenced to chassis ground. But the gate-drive signal is transformer-coupled to the FET, which is connected to a -300 VDC bus. Set up the TekScope instrument to compare the gate-drive signal at the output of the timing circuit to the signal at the gate of the FET.





METER () SCOPE			NILLIN UHII NILLIN UHII IIHUAILTIA NILLIN UHII IIHUAILTIA	TOGGLE +
SCOPE	CH 1	_	_	_
	CH 2			
	AUTO RANGE			

Setup to Test the Transistor Drive Circuit

You do not have to do anything special to take this difficult measurement. Because of the isolated channels, you can reference the channel 1 probe to chassis ground and the channel 2 probe directly to the -300 VDC bus. Channel 1 displays the gate-drive signal directly from the driver and channel 2 displays the signal as it is received by the power FET.

Going Further

The isolated channels allow you to reference a channel to AC as well as DC voltages.

- You can connect the P5102 probe reference lead to 50 Hz, 60 Hz, or 400 Hz AC power lines (up to the maximum voltage rating).
- You can connect the P6113B or P5102 probe reference lead to dynamic signals with slew rates up to 3000 V/µs (up to the maximum voltage rating).

Because you can connect to references other than ground, you can take many measurements that would otherwise require an oscilloscope with a differential input.

Measuring Instantaneous Power Dissipation in a Switching Transistor

The output transistor in a switching power supply is hotter than it should be. You are concerned about its peak power dissipation. Set up the TekScope instrument to measure the instantaneous power dissipation of the transistor using an optional current probe.



METER METER				TOGGLE +
SCOPE	CH 1	—	_	—
	CH 2	Probe Type	Current Probe	Set to 100 mV/A
	AUTO RANGE	_	_	_
	MATH	Function	$Ch1 \times Ch2$	
	CURSOR	Function	Paired	Set cursor

Setup to Measure Instantaneous Power Dissipation

Move the cursor along the power (MATH) waveform and read the instantaneous power in the cursor readout (for example, @5.63 W).

Going Further

Measure the average power dissipation in the transistor (mean value of the power waveform) with the following setup:

METER METER				TOGGLE +
SCOPE	MEAS	Select Measrmnt for MATH	Mean	—
		OK Select Measrmnt	_	

You can also display the I-V characteristic of the transistor for comparison to its safe operating area using the XY display format. Refer to page 3–14 for information on XY display format.

Triggering at a Specific Motor RPM

You need to measure the start-up current of a 3600-RPM motor at several specific speeds. A tachometer, attached to the motor, outputs a low-voltage square wave with 100 pulses per revolution. Set up the TekScope instrument to trigger at 1200 RPM so you can measure the current at that speed.

Tachometer output frequency = $\frac{1200 \text{ rev/min} \times 100 \text{ pulses/rev}}{60 \text{ s/min}}$ = 2 kHz Tachometer pulse width = $\frac{\text{period}}{2} = \frac{\frac{1}{2 \text{ kHz}}}{2} = 250 \ \mu s$ Ch 1 Ch 2 100 mV/A 00000 0 0 0 0 0 0 0 00000 $^{\circ}$ 00 \bigcirc 00 0 \bigcirc 0 0 0 Electric motor with attached tachometer @: 12.8 A Tachometer signal (Ch 1) ----Motor current (Ch 2)

THS 710 & THS 720 User Manual

Setup t	o Trigger	at 1200	RPM
---------	-----------	---------	-----

				TOGGLE +
SCOPE	CH 1	—	—	_
	CH 2	Probe Type	Current Probe	Set to 100 mV/A
	HORIZON- TAL MENU	Trigger Position	50%	_
	TRIGGER MENU	Туре	Pulse	
		Source	Ch1	
		Polarity and Width	Positive	Set width to 250µs
		Trigger When	Equal To Width	Set±5%
		Mode	Normal	_
	CURSOR	Function	Paired	Set cursor to horizontal center of graticule

Set appropriate **VOLTS/DIV** for each channel. Set **SEC/DIV** so that the sweep duration is approximately equal to the motor start-up time. When the motor starts, the 1200 RPM point is displayed at the center of the graticule. Use the cursor on channel 2 to measure the armature current at this point.

Going Further

Change the trigger pulse width setting to trigger at other specific motor RPMs to complete the test.

Using the Oscilloscope and DMM at the Same Time

You need to evaluate a VCO (voltage-controlled oscillator) circuit. You want to produce a graph of the output frequency as a function of the input control voltage. Set up the TekScope instrument to provide you with data for your graph.



METER () SCOPE				TOGGLE +
METER	AUTO RANGE	VDC	_	_
SCOPE	AUTO RANGE	_		
	MEAS	Select Measmnt for CH1	Frequency	
		OK Select Measrmnt	_	

Setup to Use DMM and Oscilloscope Together

You can use the DMM to measure the VCO control voltage and the oscilloscope to measure the output frequency at the same time.

In scope mode, the meter readout is in the upper right corner of the display. The frequency measurement is at the right side of the graticule. Both measurements update continuously as you adjust the VCO control voltage.

Going Further

You may find simultaneous DMM and oscilloscope operation useful in additional ways:

- To monitor changes in circuit operation as you adjust the power supply voltage
- To monitor the output amplitude of an optical transmitter or detector as you adjust its bias voltage

Monitoring for Power Surges and Dropouts

You are having intermittent problems with some electronic equipment that operates unattended at a remote site. You need to determine if the problem might be caused by momentary power quality problems in the electrical service to the equipment. Set up the TekScope instrument to monitor the line voltage for a week and capture any surges or dropouts that may occur.



METER (SCOPE)				TOGGLE +
METER	—	VAC	—	_
	ACQUIRE	Mode	Peak Detect	

Setup to Monitor for Power Quality Problems

You can use the DMM data logger to record measurements over an extended period of time. Set the full-scale range to 400 V using the **VOLTS/DIV** rocker. Adjust the **SEC/DIV** rocker so that the data logger horizontal scale is one day per division.

Meter measurements, which occur at a rate of approximately ten per second, are captured over an eight-day period of time.

Going Further

Use the cursors to determine when a power disturbance occurred (to within the nearest 24 minutes).



Move either cursor to the location of the power disturbance. Read the relative time in the cursor readout (for example, @ 52 h). Calculate the absolute time and date of the disturbance using the current time and this measurement.

Triggering on a Video Field

The image quality is poor on a video monitor in a closed-circuit security system that operates at a 15 kHz scan rate. Set up the TekScope instrument to display and trigger on field 1 of the video waveform coming into the monitor.



<u> </u>	-								
		·	·	•	•			•	
·		·		•••	•	•	•	•	·
			•	· · -			•		
•••	•		••••		•••	•••		· · ·	••••
••••			:	••••			••••		:.:
		•	•	•		•	•		•
				-	-				
				-	-				

Setup to Trigger on Field 1

METER () SCOPE			FILLIN UFTI FILLIN UFTI THATALISTA FILLIN UFTI FILLIN UFTI THATALISTA	TOGGLE +
SCOPE	VERTICAL MENU	Probe Type	Voltage Probe	Set to 1X
	AUTO RANGE	_	_	_
	DISPLAY	Style	Dot Accumulate	Set to 100 ms
	TRIGGER	R Type Video		_
	MENU	Trigger On	Field 1	
		Scan Rate	_	Set to 15-20 kHz

Adjust the **SEC/DIV** rocker to **2 ms/div** to display field 1 across about eight divisions. The Dot Accumulate display style simulates an analog oscilloscope display of the video signal.

Going Further

If you are testing a higher resolution video system, you can trigger on video signals with scan rates up to 65 kHz.

Triggering on a Video Line

Now you want to see one line of the staircase test pattern. Set up the TekScope instrument to display and trigger on a video line.



Setup to	Trigger	on a	Video	Line
----------	---------	------	-------	------

METER (SCOPE)				TOGGLE +
SCOPE	VERTICAL MENU	Probe Type	Voltage Probe	Set to 1X
	AUTO RANGE	_	_	_
	ACQUIRE	Mode	Peak Detect	
	TRIGGER	Туре	Video	
	MENU	Trigger On	Lines	

Adjust the **SEC/DIV** rocker to $10 \mu s/div$ to display the lines across about six divisions.

Going Further

If you need to display a specific video line, use the following method:

- **1.** Trigger on the video field that contains the line you want to display (see the *Triggering on a Video Field* application example).
- 2. Display the delayed time base. Set the delay time value to the time from the start of the field to the start of the line of interest. Refer to page 3–22 for information about using the delayed time base.
- **3.** Adjust the delayed time base **SEC/DIV** to display the video line of interest.

Reference

Introduction to Reference

This chapter contains detailed information about the operation of the THS 710 and THS 720 TekScope instruments. The topics in this chapter are arranged alphabetically by button name.

Reference Topic	Page
Acquire	3–3
Autorange	3–8
Cursor	3–11
Display	3–13
Hard copy	3–17
Hold	3–19
Horizontal controls	3–20
Measure	3–24
Meter mode	3–31
Save/Recall	3–37
Scope mode	3–40
Trigger controls	3–46
Utility	3–53
Vertical controls	3–59

ACQUIRE

Press the ACQUIRE button to set acquisition parameters independently for scope mode and meter mode.

Acquire Menu in Scope Mode

SCOPE	ACQUIRE	Mode	Sample Peak Detect Envelope Average
		Stop After	HOLD Button Only Single Acquisition Sequence
		Force Trigger	_

Key Points

Acquisition Modes. You can choose one of four acquisition modes: Sample, Peak Detect, Envelope, or Average. The next two pages describe these acquisition modes in detail.



Sample mode acquires one sample in each interval.



Peak Detect mode uses the lowest and highest samples from two consecutive intervals.

Sample. Use Sample acquisition mode for the fastest acquisition at any SEC/DIV setting. Sample mode is the default mode.

Peak Detect. Use Peak Detect acquisition mode to limit the possibility of aliasing. Also, use Peak Detect for glitch detection. You can see glitches as narrow as a 8 ns.

Peak Detect is only functional at SEC/DIV settings of 1 µs/div or slower. For SEC/DIV settings of 500 ns/div or faster, the TekScope instrument switches to Sample acquisition mode automatically.



Envelope and Average. Use Envelope acquisition mode to capture variations of a signal over a longer period of time. Use Average acquisition mode to reduce random or uncorrelated noise in the signal you want to display.

The $^+/-$ rocker sets a specific number of acquisitions (N) to include in the enveloped or averaged waveform.

- The enveloped waveform clears and then starts over after N acquisitions.
- The averaged waveform is a running average over N acquisitions.
- If you select Stop After Single Acquisition Sequence, an envelope or average acquisition stops after N acquisitions.
If you probe a noisy square wave signal that contains intermittent, narrow glitches, the waveform displayed will vary depending on the acquisition mode you choose.



Single Acquisition Sequence. The content of a single acquisition sequence depends on the acquisition mode.

Acquisition Mode	Single Acquisition Sequence
Sample or Peak Detect	One acquisition of each displayed channel
Envelope or Average	N acquisitions of each displayed channel (N is user adjustable)

Acquire Menu in Meter Mode

METER ACQUIRE		Mode	Sample Peak Detect Average
		Rel Δ	On (Reset ∆) Off

Key Points

Acquisition Modes. The data logger compresses a sequence of meter measurements into a point and then plots a series of those points to form a graph. The acquisition mode determines how the graph is calculated:

- For each point, Sample displays the first meter measurement from the sequence.
- Peak Detect displays a column representing the maximum and minimum meter measurements during the sequence.
- Average displays the average of all meter measurements during the sequence.

Rel Δ Measurements. Rel Δ acquires a new baseline value for subsequent DMM measurements. Use Rel Δ to store the current DMM value and then measure the relative change from that value. When you turn off Rel Δ , the baseline value resets to zero.

AUTORANGE

Autorange automatically adjusts setup values to track a signal. If the signal changes, the setup continues to change to track the signal. Autorange works independently in scope and meter modes.

The following controls are preset when you first select the autorange function.

Scope Mode	Meter Mode
Acquire mode: Sample	none
Stop acquire after: HOLD button only	
Vertical coupling: DC (if GND was selected)	
Bandwidth: Full	
Invert: Off	
Horizontal position: Centered	
Horizontal magnification: Off	
Trigger type: Edge	
Trigger source: Lowest numbered channel displayed	
Trigger coupling: DC	
Trigger slope: Positive	
Trigger holdoff: Minimum	
Display style: Vectors	
Display format: YT	

These conditions start an autorange cycle.

Scope Mode	Meter Mode
Too many or too few waveform periods for a clear display of the lower-numbered channel	DMM reading exceeds ±3600 counts or falls below ±330 counts
Waveform amplitude too large or too small compared to full screen if only one channel is displayed	
Waveform amplitude too large or too small compared to half screen if two channels are displayed	

Autorange adjusts these controls.

Scope Mode	Meter Mode
Vertical VOLTS/DIV adjusted	Range adjusted
Horizontal SEC/DIV adjusted	
Trigger level set to 50%	

These control changes turn off autorange.

Scope Mode	Meter Mode
Change to Stop After Single Acquisi- tion Sequence	Change range (VOLTS/DIV)
Change VOLTS/DIV	
Change SEC/DIV	
Change trigger type	
Change trigger level	
Change trigger coupling	
Change trigger holdoff	
Change display format to XY	
Change display style	

CURSOR

Press the CURSOR button to display the cursor menu. In scope mode, cursors operate on the 2500-point record of the selected waveform. In meter mode, they operate on the 250-point data logger plot.

SCOPE	CURSOR	Function	Off H Bars V Bars Paired
		Time Units	Seconds 1/seconds (Hz)
METER	CURSOR	Function	Off H Bars V Bars Paired

Key Points

Cursor Movement. Use the $^+/-$ rocker to move the active cursor. Press the TOGGLE button to change which cursor is active.

Fine Cursor Movement. If you first press MAG, you can set a cursor to any point in the 2500-point oscilloscope waveform.

Cursor Functions. H Bars measure voltage. V Bars measure time or frequency. Paired measures both voltage and time or voltage and frequency.



Paired cursors

@ Readout. For V Bar cursors, the readout after the @ symbol indicates the location of the active cursor relative to the trigger point. For H Bars or Paired cursors, it indicates the location relative to zero volts.

DISPLAY

Press the DISPLAY button to choose how waveforms are presented and to change the appearance of the display.

Display Menu in Scope Mode

METER METER				TOGGLE +
SCOPE	DISPLAY	Style	Vectors Dots	_
			Vector Accumu- late Dot Accumulate	Set accu- mulate time
		Display Contrast	_	Set contrast
		Display 'T' Trig Pt	On Off	_
		Graticule	Full Grid Cross Hair Frame	
		Format	YT XY	

Key Points

Display Style. Choose one of the following waveform display styles:

- Vectors fills the space between adjacent sample points in the display. Widely spaced points are filled using (sin x)/x interpolation.
- Dots displays only the individual sample points.
- Vector Accumulate adds persistence to the vector display. Use the ⁺/- rocker to set the accumulate time.
- Dot Accumulate adds persistence to the dot display. Use the +/- rocker to set the accumulate time.

NOTE. Vector Accumulate and Dot Accumulate are display functions only. When you change most control settings, the accumulated data is cleared. Accumulated waveforms cannot be saved.

Display 'T' at Trig Pt. The trigger point on the waveform is marked with the T symbol. The T symbol can be turned on or off.

XY Format. Choose XY display format when you want to display channel 1 in the horizontal axis and channel 2 in the vertical axis. The controls operate as follows:

- The channel 1 VOLTS/DIV and vertical POSITION controls now set the horizontal scale and position.
- The channel 2 VOLTS/DIV and vertical POSITION controls continue to set vertical scale and position.
- The SEC/DIV and horizontal POSITION controls affect the time base and the portion of the waveform that is displayed.



NOTE. The above XY-display example shows the I-V characteristic of a switching power MOSFET. The current waveform, displayed in the vertical axis, is measured using a Tektronix A6302 current probe and AM503B current-probe amplifier.

The following functions do not work in XY display format:

- Ref or Math waveforms
- Cursors
- Horizontal MAG
- Autorange (resets display format to YT)

Display Menu in Meter Mode

METER METER				TOGGLE +
METER	DISPLAY	Style	Thin Thick	—
		Display Contrast	—	Set contrast
		Graticule	Full Grid Cross Hair Frame	_

Key Points

Data Logger Display Style. For better visibility, choose Thick for a three-pixel-high data logger plot. The default (Thin) is one pixel high.

HARD COPY

You can print a hard copy of the display if a printer is connected and properly configured. Press the HARD COPY button to start printing. If you do not want menus to show, press CLEAR MENU before you press HARD COPY. You cannot change instrument settings while the hard copy is printing.

Connecting a Printer

Use the RS-232 cable to connect the printer to the RS-232 port on the side of the TekScope instrument. The RS-232 accessory kit includes an adapter for 9-pin RS-232 connectors.

- Refer to RS-232 System on page 3–55 for information about setting the TekScope instrument for RS-232 communication with your printer.
- Refer to the user manual for your printer for information about setting its baud rate and any other required parameters.

TekScope instrument



Setting Up to Print

Perform the following steps to choose the printer and page layout:

METER METER			
SCOPE or UTILIT METER	UTILITY	System	Hard Copy
		Layout	Landscape Portrait
		Format	three pages of formats
		Select Page	_
		OK Select Format	_

The following printer and file formats are supported:

- BMP (Microsoft Windows file format)
- Deskjet (high resolution printer format)
- DPU 411/II, HC 411 (thermal printer format)
- DPU 412 (thermal printer format)
- EPS Image (encapsulated postscript image file format)
- Epson (9-pin and 24-pin dot matrix printer format)
- Interleaf .img (image object file format)
- Laserjet (laser printer format)
- PCX (PC Paintbrush monochrome image file format)
- Thinkjet (inkjet printer format)
- TIFF (tag image file format)

HOLD

Press the HOLD (RUN/STOP) button to stop and start data acquisition. Because scope mode and meter mode have independent acquisition states, the HOLD button operates independently for these two modes.



Hold Function in Scope Mode

In scope mode, the function of the HOLD button depends on the Stop After selection in the acquire menu.

Acquire Menu Setting	Functions of HOLD Button
Stop After HOLD Button Only	First press stops waveform acquisi- tion.
	Second press restarts waveform acquisition.
Stop After Single Acquisition Sequence	Any press starts a new acquisition sequence. The acquisition sequence stops automatically.



Hold Function in Meter Mode

In meter mode, pressing the HOLD button once freezes the meter reading, the measurement statistics (MIN, MAX, and AVG), and the data logger display.

A second press of the HOLD button resets and restarts the data logger display and measurement statistics, and then restarts the meter readings.

HORIZONTAL Controls

You can use the horizontal controls to change the time base, horizontal position, and horizontal magnification of waveforms.



Horizontal Operations in Scope Mode

METER () SCOPE			NILINUFI NILINUFI IIHU AILINA NILINUFI IIHU AILINA	TOGGLE +
SCOPE	HORIZON-	Time Base	Main	—
	TAL MENU	Delayed Runs After Main	Set delay time	
		Trigger Position	Set to 10% Set to 50% Set to 90%	_
			% Pretrigger	Set %

Key Points

SEC/DIV Rocker. If waveform acquisition is stopped (using the HOLD button), changes you make to the time base have no effect until you restart acquisition.

Roll Mode Display. To obtain a rolling display similar to a strip-chart recorder, select Auto trigger mode and set SEC/DIV to 500 ms/div or slower.

POSITION Rocker. You can position each of the two reference waveforms (Ref A and Ref B) independently of the three live waveforms (Ch 1, Ch 2, and Math). Or you can set the horizontal position of all waveforms track each other. Refer to *Ref A or Ref B Vertical Menu* on page 3–62 for information about this capability.

MAG Button. To switch between the normal and magnified displays, press the MAG button.

- Normal display compresses the 2500-point waveform by a factor of ten to form 250 horizontal points in the display.
- Magnified display expands the horizontal scale by a factor of ten and displays one waveform point per pixel.
- Use the POSITION rocker to choose the section of the waveform to magnify. The horizontal-position indicator shows you the location of the magnified segment in the full waveform record.



Readout. The waveform readout shows the horizontal scale factor below the graticule. Page 3–40 shows the location of this readout.

Time Base. Choose the Main or Delayed time base. The delayed time base runs at the preset delay time after the trigger event for the main time base. Use the $^{+}/^{-}$ rocker to set the delay time.



Trigger Position. Choose the amount of pretrigger:

- Set to 10% places the trigger point near the beginning of the waveform record
- Set to 50% places the trigger point at the center of the waveform record
- Set to 90% places the trigger point near the end of the waveform record
- You can also set any amount of pretrigger (0% to 100%) with the ⁺/- rocker

Horizontal Operations in Meter Mode

METER	HORIZONTAL MENU	_	_

Key Points

SEC/DIV Rocker. To adjust the scroll speed of the data logger plot, use the SEC/DIV rocker. If you change the scroll speed, data in the logger display is erased.

Other Controls. The POSITION rocker and MAG button have no effect in meter mode.

MEAS

You can use the MEAS button to access the automatic measurement capability of the TekScope instrument. In scope mode, the instrument measures the 2500-point, selected waveform. In meter mode, the measurements take the form of statistics, which are calculated from successive meter readings.

METER Scope				TOGGLE +
SCOPE	MEAS	Select Meastmnt	six pages of measurements	—
		Select Page	_	
		Remove Measrmnt	Measrmnt	Choose mea- surement
			All Measrmnts	_
		High-Low Setup	Histogram Min-Max	
		OK Select Measrmnt	_	
		OK Remove Measrmnt		

Measurements in Scope Mode

Key Points

Choosing Measurements. You can perform up to four automatic measurements on the selected waveform and display them along the right side of the graticule. The table beginning on page 3–26 describes the scope-mode measurements in detail.



High-Low Setup. The TekScope instrument determines the 10%, 50%, or 90% levels of the selected waveform and then uses them to calculate the measurements. You can choose the method used to determine these levels:

- Histogram sets the values statistically; it finds the most common value either above or below the midpoint (depending on whether it is defining the high or low reference level). Since this statistical approach ignores short-term aberrations (overshoot, ringing, noise), histogram is the best method for measuring digital waveforms and pulses.
- Min-max uses the highest and lowest values of the waveform record. This method is best for measuring waveforms that have no large, flat portions at a common value, such as sine waves and triangle waves.

Scope-Mode Measurement Definitions

Name		Definition
<u></u>	Ampl	Measured over the entire waveform.
		Amplitude = High (100%) – Low (0%)
-MN-L	BrstW	The duration of a burst. Measured over the entire waveform.
	cMean	The arithmetic mean over the first cycle in the waveform.
2/2	cRMS	The true Root Mean Square voltage over the first cycle in the waveform.
f	Fall	Time that the falling edge of the first pulse in the waveform takes to fall from 90% to 10% of its amplitude.
_* _*	Freq	Reciprocal of the period of the first cycle in the waveform. Measured in Hertz (Hz).
	High	The value used as 100%. Calculated using either the min/max or the histogram method. Measured over the entire waveform.
<u>. [.] _</u>	Low	The value used as 0%. Calculated using either the min/max or the histogram method. Measured over the entire waveform.

Scope-Mode Measurement Definitions (Cont.)

Name	Definition
	The maximum amplitude. The most positive peak voltage measured over the entire waveform.
John Mean	The arithmetic mean over the entire waveform.
<u></u> Min	The minimum amplitude. The most negative peak voltage measured over the entire waveform.
_ <u></u> *_∗ –Duty	Measurement of the first cycle in the waveform. Negative Duty Cycle = $\frac{Negative Width}{Period} \times 100\%$
–Over	Measured over the entire waveform. Negative Overshoot = $\frac{Low-Min}{Amplitude} \times 100\%$
–Width	Measurement of the first negative pulse in the waveform. The time between the 50% amplitude points.
Pk-Pk	Measured over the entire waveform. <i>Amplitude = Max – Min</i>

Scope-Mode Measurement Definitions (Cont.)

Name	Definition
_ <u>⊀</u> ⊦ Period	Time it takes for the first complete signal cycle to complete in the waveform. Measured in seconds.
_⊀ŧ厂 +Duty	Measurement of the first cycle in the waveform. Positive Duty Cycle = $\frac{Positive Width}{Period} \times 100\%$
+Over	Measured over the entire waveform. Positive Overshoot = $\frac{Max-High}{Amplitude} \times 100\%$
∗ ∗ +Width	Measurement of the first positive pulse in the waveform. The time between the 50% amplitude points.
Rise	Time that the leading edge of the first pulse in the waveform takes to rise from 10% to 90% of its amplitude.
my RMS	The true Root Mean Square voltage over the entire waveform.

Measurements in Meter Mode

METER () () () () () () () () () ()				TOGGLE +
METER	MEAS	Select Statistic for DMM	Max Avg Min Rel Δ Max – Min	_
		Select Page	—	
		Remove Statistic	Statistic	Choose statistic
			All Statistics	—
		Beep New Max-Min	On Off	
		OK Select Statistic	_	
		OK Remove Statistic		

Key Points

Data Included in Statistics. The statistics are calculated over all meter readings since the last reset. The statistics can represent data that is no longer displayed in the data logger plot.

Readout. The statistic readouts appear in the upper-right corner of the display. See page 3–31 for the location.

Statistic	Definition
Мах	The maximum value of all meter readings since the last reset.
Avg	The arithmetic average of all meter readings since the last reset.
Min	The minimum value of all meter readings since the last reset.
Δ0	The baseline value used in the Rel Δ calculation. This value is updated only when the Rel Δ function is turned on or off.
Max-Min	The difference between the maximum and minimum meter readings since the last reset.

Choosing Statistics. Choose up to three from the following statistics.

Reset Conditions. The calculated statistics are reset to zero if you make any of the following control changes:

- Turn off the HOLD function
- Change the measurement function (from VAC to VDC, for example)
- Change the probe scale factor
- Change the $Rel\Delta$ value
- Change the data logger scroll speed

Beep New Max/Min. You can turn on a beep that sounds any time the TekScope instrument updates the Max or Min statistic.

METER Mode



Press the METER button to enter meter mode. The meter reading and statistics update about three times per second.

METER	METER	VAC	—
		VDC	—
		Ω	_
		(continuity)	—
		-⊳+- (diode)	_

The meter-mode display, shown below, is divided into three sections. The next two pages identify the content of each section in detail.



Status Line

The status line across the top of the display contains acquisition information. The overrange indicator warns when an overvoltage is applied to the input.

Acquisitio	n readout	Overrange indicator
TekRun	Pk Detect	Over Range

The table below shows examples of the acquisition readout.

Acquisition Readout	Explanation	
AUTO Rânge	Acquisition running with autorange function on (AUTO	
Run	RANGE) of oil (Ruil), of acquisition is stopped (noid)	
Hold		
Data: 7	Saved data is recalled for display (from location 7) while acquisition runs in the background	
Sample	Acquisition mode for data logger	

Graticule Area

The graticule area contains the data logger plot, the bar graph, and their scale markers.



Measurement Readout Area

The area above and to the right of the graticule contains the current meter reading, statistic readouts, and cursor readouts.



Data Logger Display

The data logger records meter measurements over a period of time, creating a plot similar to a strip-chart recorder. You can set the time span of the plot from four minutes to eight days.

The data logger plot scrolls to the left. The most recent data is always at the right end of the graticule. The oldest data disappears off the left side of the graticule and is erased.



Reset. The TekScope instrument erases the waveform in the data logger plot if you make any of the following control changes:

- Turn off the HOLD function
- Change the measurement function (from VAC to VDC, for example)
- Change the probe scale factor
- Change the $\text{Rel}\Delta$ value
- Change the data logger scroll speed

Zero Level. If you select the VDC meter function, the zero level is located at the horizontal centerline of the graticule; for all other meter functions, the zero level is the bottom of the graticule.

Reference Level. If the Rel Δ function is turned on, the Rel Δ baseline value is marked along the left side of the graticule.

VOLTS/DIV Rocker. Use the VOLTS/DIV rocker to set the meter range and control the vertical scale of the data logger plot. If you change the vertical scale while the data logger is running, you will cause a vertical discontinuity; old data is not rescaled to match the current setting. However, if you change the vertical scale while the HOLD function is active, the entire waveform is scaled to match the current setting.

SEC/DIV Rocker. Use the SEC/DIV rocker to control the scroll speed of the data logger plot. When you change the scroll speed, existing data in the data logger plot is erased.

Bar Graph

The rapid update rate of the bar graph simulates an analog meter movement. The bar graph is displayed just to the right of the data logger display and uses the vertical axis of the data logger display as its scale. The bar graph extends from zero (or the Rel Δ baseline value) to the current meter measurement. It tracks the meter measurement rate and updates about ten times per second.

In addition, the bar graph contains two hollow segments that represent the current values of the Min and Max statistics.



For More Information

Many of the other sections in this chapter contain additional information about meter-mode menus and operation. Please refer to those sections, which are arranged alphabetically by button name.

SAVE/RECALL

Press the SAVE/RECALL button to save or recall any of the following:

- Setups
- Oscillscope waveforms
- DMM Data

Save/Recall Menu in Scope Mode

METER () SCOPE				TOGGLE +
SCOPE	SAVE/ RECALL	Save Current Setup	To Setup	Choose setup location
		Recall Saved Setup	Recall Factory Setup	_
			Recall Setup	Choose setup location
		Save selected wfm	To Waveform	Choose wave- form location
		Recall Saved Waveform	Load REFA From Wfrm	
			Load REFB From Wfrm	
		OK Save Setup	_	_
		OK Recall Setup		
		OK Recall Factory		
		OK Save Waveform		
		OK Recall Waveform		

Key Points

Saving and Recalling Setups. Whether you save a setup in scope mode or meter mode, the TekScope instrument stores its complete setup in nonvolatile memory. When you recall the setup, you will be in the mode from which the setup was saved.

Recalling the Factory Setup. You can recall the Factory Setup to initialize the TekScope instrument to a known setup. *Appendix B* describes the Factory Setup in detail.

Saving a Waveform. Push the CH 1, CH 2, or MATH button to choose the waveform to save. Waveform position and scale factors are saved with each waveform.

Recalling a Waveform. Recall a saved waveform into either Ref A or Ref B for display. When you recall a saved waveform, the recalled waveform overwrites the previous Ref A or Ref B waveform.

Saving and Displaying a Waveform in One Step. Using a vertical menu, you can save a waveform and keep it for display at the same time. Refer to *Ref A or Ref B Vertical Menu* on page 3–62 for information about this capability.

Save/Recall Menu in Meter Mode

METER (SCOPE)				TOGGLE +
METER	SAVE/ RECALL	Save Current Setup	To Setup	Choose setup location
		Recall Saved Setup	Recall Factory Setup	_
			Recall Setup	Choose setup location
	Save Data Recall	Save DMM Data	To Data	Choose DMM data location
		Recall DMM	Recall Data	
		Data	Clear Data From Screen	
		OK Save Setup	_	_
		OK Recall Factory		
		OK Save Data		
		OK Recall Data		
		OK Clear Data		

Key Points

Saving and Recalling Setups. Whether you save a setup in scope mode or meter mode, the TekScope instrument stores its complete setup in nonvolatile memory.

Saving DMM Data. Saving DMM data saves the meter mode, meter range, current DMM reading, statistics, and data logger plot.

SCOPE Mode

Press the SCOPE button to enter scope mode. If already in scope mode, pressing this button has no effect.

The scope-mode display, shown below, is divided into four sections. The next five pages identify the content of each section in detail.



Waveform readout lines

Status Line

The status line across the top of the display contains acquisition and trigger information.



The table below shows acquisition readout examples that are displayed when acquisition is running. When you press HOLD to stop acquisition, the readout shows the number of waveforms acquired since acquisition was last stopped.

Acquisition Readout	Explanation
AUTO RANGE	Acquisition running with autorange function on (AUTO
Run:	
25MS/s	Current sample rate
Pk Detect	Acquisition mode

The table below shows the meaning of the trigger status indicators.

Trigger Status	Explanation
Auto	Free-running in Auto trigger mode
Trig?	Waiting for trigger in Normal trigger mode
PrTrig	Acquiring new pre-trigger data
The table below shows examples of additional items that are sometimes displayed in the status line.

Additional Items	s Explanation	
🔄 -3.253 VDC	DMM icon and current DMM reading	
🕂 Delay: 🖃 1.014µs	Parameter and its current value (only when the +/– rocker is assigned to a parameter)	

Graticule Area

The graticule area contains waveforms and position indicators.



Waveform Readout Lines

The readout lines below the graticule contain specific information about displayed waveforms. The top line displays vertical readout for channel 1 and channel 2. The bottom line displays readout for Ref A, Ref B, or Math, whichever waveform was selected last.



The table below shows examples of the vertical readout symbols.

Vertical Readout Symbols	Explanation
612	Selected waveform
Ch 1	Unselected waveform
÷	Inverted channel
ក	GND coupling
	AC coupling
B _W	Bandwidth limit on
B-8	Reference waveform Ref B recalled from waveform storage location eight

The waveform readout lines also show time base and trigger information.



The table below shows examples of the trigger information.

Trigger Information	Explanation		
Ch 1	Trigger source		
~	Edge trigger slope		
-148mV	Trigger level		
л	Pulse trigger polarity		
>	Pulse trigger condition		
990ns	Pulse trigger width		
Field 2	Video trigger condition		

Measurement Readout Area

The area to the right of the graticule contains cursor and measurement readouts. If a measurement qualifier appears with a measurement result, the signal may be insufficient to take an accurate measurement.



For More Information

Many of the other sections in this chapter contain additional information about scope-mode menus and operation. Please refer to those sections, which are arranged alphabetically by button name.

TRIGGER Controls

Triggering is an oscilloscope-only function; the trigger controls, shown below, have no effect in meter mode.



The trigger types are the following:

- Edge triggers on the rising or falling edge of the input signal (see page 3–48).
- Pulse triggers on specific events that you can qualify by time (see page 3–50).
- Video triggers on field 1, field 2, or a specific line of an NTSC or PAL standard video signal (see page 3–52).

Use the leftmost button in the trigger menu to choose the trigger type. The remaining items in the trigger menu depend on the trigger type you choose.

	Edge		Pulse	Video	
Source	Choose Ch 1 or Ch 3	Source	Choose Ch 1 or Ch 2	Source	Choose Ch 1 or Ch 2
Coupling	DC DC	& width	Positive 1	igger on	Field 1 (Interlaced)
	Reject	olarity			(Interlaced)
	LF Manager Reject	ď			Any Field (Non-interlaced)
	Noise Rej (DC Low Sensitivity)		¹ Set pulse width with ⁺ /– rocker		Lines
Slope	Positive	. when	Less Than Width →	n rate	15 kHz to 20 kHz
	Negative 🔨	Trigger	Greater Than Width	Sca	20 kHz to 25 kHz
			Equal To _j−†,j⊥ Width ²		25 kHz to 35 kHz
			Not Equal To Width ²		35 kHz to 50 kHz
			² Set tolerance		50 kHz to 65 kHz

Edge Trigger

Use Edge triggering to trigger on the rising or falling edge of the input signal at the trigger threshold.

METER C SCOPE			NITLINUFII NITLINUFII IIHUMIJIM NITLINUFII IIHUMIJIM	TOGGLE +
SCOPE	TRIGGER	Туре	Edge	—
	MENU	Source	Ch1 Ch2	
		Coupling	DC HF Reject LF Reject Noise Reject	
		Slope	/ (rising edge) \ (falling edge)	
		Mode & Holdoff	Auto Normal	Set holdoff

Key Points

Normal and Auto Mode. Use Normal trigger mode when you want the oscilloscope to trigger only on a valid trigger. Use Auto trigger mode when you want the acquisition to free-run in the absence of a valid trigger event. Also, choose Auto when you want an untriggered, rolling waveform at 500 ms/div or slower time base settings.

Holdoff. You can use holdoff to help stabilize the display of nonperiodic waveforms. After you press the Mode & Holdoff menu button, use the $^{+}/^{-}$ rocker to set the holdoff time from 500 ns to 10 s.

Holdoff begins when the TekScope instrument recognizes a trigger event and disables the trigger system until acquisition is complete. The trigger system remains disabled during the holdoff time that follows each acquisition.



Triggers are not recognized during holdoff time.

NOTE. For best results, choose Normal trigger mode when using long holdoff settings (10 ms or greater).

Pulse Trigger

Use Pulse triggering to isolate and display specific events that you can qualify by time.

METER SCOPE			NICLINUFII NICLINUFII IHUUILIJIA NICLINUFII IHUUILIJIA	TOGGLE +
SCOPE	TRIGGER	Туре	Pulse	—
	MENU	Source	Ch1 Ch2	
		Polarity & Width	Positive Negative	Set width
		Trigger When	Less Than Width	_
			Greater Than Width	
			Equal To Width	Set±%
		Not Equal To Width		
		Mode & Holdoff	Auto Normal	_

Key Points

Trigger Conditions. You can trigger on the following conditions:

- Less Than Width triggers on a pulse width that is less than the time threshold.
- Greater Than Width triggers on a pulse width that is greater than the time threshold.

- Equal To Width triggers on a pulse that matches the set pulse width within a given tolerance. Use the ⁺/– rocker to set the tolerance in percent. For example, if the pulse width is set to 1 μs and the tolerance is set to ±20%, triggering occurs only on pulse widths in the range from 800 ns to 1.2 μs.
- Not Equal To Width triggers on a pulse that does not match the set pulse width and tolerance. Use the ⁺/– rocker to set the tolerance.

Trigger Location. The T symbol shows where triggering occurs for the four trigger conditions.



Video Trigger

Choose video triggering to trigger on field 1, field 2, or the lines of an NTSC, PAL, or SECAM standard video signal. Also, you can trigger on nonstandard video signals with scan rates up to 65 kHz.

METER (SCOPE)				TOGGLE + -
SCOPE	TRIGGER	Туре	Video	-
MENU	Source	Ch1 Ch2		
		Trigger On	Field 1	
			Field 2	
			Any Field	
			Lines	
		Scan Rate	_	Set scan rate
		Mode & Holdoff	Auto Normal	Set holdoff

Key Points

Sync Pulses. When you choose Video, the trigger always occurs on negative-going sync pulses. If your video signal has positive-going sync pulses, invert the signal using the vertical menu. Refer to *VERTICAL Controls* on page 3–59 for information about inverting a signal.

UTILITY

The following are examples of what you can do with each of the six branches in the Utility menu:

- Use Config to display the firmware version.
- Use Hard Copy to set up hard copy parameters. Refer to *HARD COPY* on page 3–17 for information about setting up and printing a hard copy.
- Use RS-232 to set up for remote communication.
- Use Misc to set the standby or backlight timeout.
- Use Cal to compensate the signal path.
- Use Diag to run internal diagnostic routines.

Press the UTILITY button to display the utility menu. You can access the same utility menu from scope mode or meter mode. Then, use the left-most button in the utility menu to choose the branch. The remaining items in the utility menu may change depending on the branch you choose.

Config System

SCOPE or	UTILITY	System	Config
METER		Tek Secure Erase Memory	_
		Version	
		OK Erase Setup/Data	

Key Points

Tek Secure. If you have acquired confidential data, you may want to execute Tek Secure before you return the TekScope instrument to general use. Executing Tek Secure accomplishes the following tasks:

- Replaces all waveforms (oscilloscope and data logger) in all reference memories with zero sample values
- Replaces the current front-panel setup and all stored setups with the factory setup
- Calculates the checksums of all waveform memory and setup memory locations to verify successful completion of waveform and setup erasure
- Displays a confirmation or warning message if the checksum calculation is successful or unsuccessful

RS-232 System

METER () SCOPE				TOGGLE +
SCOPE or	UTILITY	System	RS-232	—
METER		Baud Rate	_	Choose rate
		Flagging	Hard Flagging	On
			Soft Flagging	Off
		Misc	EOL	CR LF CR/LF LF/CR
			Parity	None Even Odd
			Stop Bits	1 2
			Delay	Set delay
		Set RS232 Parameters to Defaults	_	_

Key Points

RS-232 Troubleshooting. If you are having RS-232 communication difficulties, try the following remedies:

- Verify that the RS-232 cable is connected to the correct port on your computer or hard copy device.
- Reset the RS-232 parameters to defaults and then set the baud rate to match the computer or hard copy device. The default settings (except baud rate) are standard on most computers and hard copy devices.

Misc System

METER Cope				TOGGLE +
SCOPE or UTILITY METER		System	Misc	_
		Power Off Time-Out	—	Set time
		Backlight Time- Out	_	Set time

Key Points

Power Off Time-Out. Use this feature to automatically turn off the TekScope instrument if it is not being used. Use the $^+/^-$ rocker to set the power off time-out delay from one minute to 15 minutes or to ∞ (time-out off).

Power off time-out operates only when you use battery power.

Backlight Time-Out. Press this button to adjust the backlight time-out delay. This feature automatically turns the backlight off after a period of time if left unattended. Use the $^{+}/_{-}$ rocker to set the backlight time-out delay from one minute to 15 minutes or to ∞ (time-out off).

Backlight time-out operates only when you use battery power.

Cal System

SCOPE or	UTILITY	System	Cal
METER		Signal Path	—
		Factory Scope	
		Factory DMM	
		OK Compensate Signal Path	
		OK Factory Cal Scope	
		OK Factory Cal DMM	

Key Points

Signal Path Compensation. Signal path compensation optimizes the oscilloscope accuracy for the current ambient temperature. For maximum accuracy, recompensate the signal path if the ambient temperature changes by 5° C or more.

To compensate the signal path, disconnect any probes or cables from the channel 1 and channel 2 input BNC connectors. Then, press the **OK Compensate Signal Path** button to confirm that you are ready to proceed.

Factory Scope and Factory DMM. Service personnel use these functions to calibrate the oscillscope and DMM internal voltage references. Refer to your Tektronix field office or representative for assistance with these processes.

Diag System

METER METER			
SCOPE or	UTILITY	System	Diag
METER		Execute	_
		Loop	Once Always Until Fail
		Error Log	_
		OK Run Test	
		OK Display Log	

Key Points

Starting Diagnostics. To execute the built-in diagnostic routines, disconnect all cables, probes, or leads from the oscilloscope and DMM inputs, and then press the **OK Run Test** button.

Stopping Diagnostics. Choose how you want the diagnostic routines to execute:

- Loop Once runs all diagnostic routines one time and then stops.
- Loop Always runs the diagnostic routines continuously. Press
 HOLD and then CLEAR MENU to resume normal operation.
- Loop Until Fail runs the diagnostic routines until the TekScope instrument fails a test or until you cycle the power.

Error Log. The error log contains summary data gathered over the life of the TekScope instrument and descriptions of the last 100 errors encountered. The last error in the list is the most recent. Press the ⁺/– rocker to display subsequent pages of the error log.

VERTICAL Controls

You can use the vertical controls to display waveforms, adjust vertical scale and position, and set input parameters.



Vertical Operations in Scope Mode

All vertical operations affect the selected waveform. Press the CH 1, CH 2, MATH, REF A, or REF B button to select that waveform.

To remove a waveform from the display, select the waveform and then press the WAVEFORM OFF button.

Press the MENU button to display the vertical menu. The contents of the vertical menu depend on which waveform is selected.

Channel 1 or Channel 2 Vertical Menu

The vertical menu contains the following items when channel 1 or channel 2 is the selected waveform.

METER SCOPE			NILINUFII HILINUFII HUUIIIII HUUIIIII HUUIIIII HUUIIIII	TOGGLE +
SCOPE	VERTICAL MENU	Coupling	DC AC GND	—
		Invert	Invert Off Invert On	
		Bandwidth	Full Bandwidth 20 MHz	
		Position	_	
		Probe Type	Current Probe	Set conversion factor
			Voltage Probe	Set probe attenuation

Key Points

GND Coupling. Use GND coupling to display a zero-volt waveform. When you use GND coupling, the input BNC connector is disconnected from internal circuits. Internally, the channel input and its reference are connected to create a zero-volt reference level.

VOLTS/DIV Rocker. Use the VOLTS/DIV rocker to change the vertical sensitivity when acquisition is running. When acquisition is stopped, the rocker scales the waveform vertically.

Math Vertical Menu

The vertical menu contains the following items when Math is the selected waveform.

METER SCOPE				TOGGLE +
SCOPE	VERTICAL MENU	Function	$\begin{array}{l} Ch1 + Ch2\\ Ch1 - Ch2\\ Ch2 - Ch1\\ Ch1 \times Ch2 \end{array}$	

Key Points

Math Waveform Units. The waveform math function recognizes the following combinations of units.

Channel 1 Unit	Channel 2 Unit	Math Operation	Resulting Math Unit
V	V	+ or –	V
A	А	+ 0r –	А
V	V	×	VV
A	А	×	AA
V	А	×	W
А	V	×	W

VOLTS/DIV Rocker. Use the VOLTS/DIV rocker to scale the waveform vertically. The rocker does not affect channel 1 or channel 2 sensitivity.

Ref A or Ref B Vertical Menu

The vertical menu contains the following items when Ref A or Ref B is the selected waveform.

METER Scope			NITLINUEII NITLINUEII NITLINUEII NITLINUEII NITLINUEII	TOGGLE +
SCOPE VERTICAL MENU	VERTICAL	Save Ch1	To Waveform	Choose wave- form location
	MENU	Save Ch2		
		Save MATH		
	Horizontal Position	Lock Ind	_	
		OK Save Waveform	_	

Key Points

Saving and Displaying a Waveform in One Step. You can copy a live waveform from one of the save sources shown above (Ch 1, Ch 2, or Math) into Ref A or Ref B for display and also into the nonvolatile storage location you choose with the $^+/-$ rocker.

Horizontal Position. Choose locked or independent horizontal position control for the reference waveforms.

- Choose Lock to position all displayed waveforms as a group.
- Choose Ind to position each reference waveform independently. The live waveforms (Ch 1, Ch 2, Math) still position as a group.

Vertical Operations in Meter Mode

METER (SCOPE)				TOGGLE +
METER	VERTICAL MENU	Noise Reject	None 60 Hz 50 Hz	_
		Probe Type	Current Probe	Set conversion factor
			Voltage Probe	—

Key Points

Noise Reject. You can improve the repeatability of AC measurements in the presence of 50 Hz or 60 Hz noise by choosing noise rejection.

VOLTS/DIV Rocker. Use the VOLTS/DIV rocker to change the meter range and the vertical scale of the data logger plot.

Appendices

Appendix A: Specifications

This appendix contains the oscilloscope, DMM, and general specifications for the THS 710 and THS 720 TekScope instruments. All specifications are guaranteed unless noted "typical." Specifications that are marked with the ν symbol are checked in *Appendix D: Performance Verification*.

All specifications apply to both the THS 710 and THS 720 unless noted otherwise. All specifications assume horizontal MAG is off, unless noted otherwise. To meet specifications, two conditions must first be met:

- The TekScope instrument must have been operating continuously for ten minutes within the operating temperature range specified.
- You must perform the Compensate Signal Path operation, accessible through the utility menu, if the operating temperature changes by more than 5° C.

Oscilloscope Specifications

Acquisition			
Acquisition Modes	Sample (Normal), Peak detect, Envelope, and Average		
Acquisition Rate, typical	Up to 25 waveforms per second (2 channels, sample acquisition mode, MAG on, no measurements)		
Single Sequence	Acquisition Mode	Acquisition Stops After	
	Sample, Peak Detect	Single acquisition, one or two channels simultaneously	
	Average, Envelope	N acquisitions, one or two channels simultaneously, N is settable from 2 to 256 or ∞	

Inputs			
Input Coupling	DC, AC, or GND		
Input Impedance, DC Coupled	1 M Ω ±1% in parallel with 25 pF ±2 pF		
Maximum Voltage	Overvoltage Category	Maximum Voltage	
Between Signal and Common at Input BNC	CAT II Environment	300 V _{RMS}	
	CAT III Environment	150 V _{RMS}	
	For steady-state sinusoidal waveforms, derate at 20 dB/decade above 100 kHz to 13 V _{pk} at 3 MHz and above. Also, refer to Overvoltage Category description on page A–15.		
Maximum Voltage Between Common and Earth Ground at BNC	30 V _{RMS} , 42.4 V _{pk}		
Channel-to-Channel Common Mode Rejec- tion, typical	100:1 at frequencies ≤50 MHz, measured on MATH Ch1 – Ch2 waveform, with test signal applied between signal and common of both channels, and with the same VOLTS/DIV and coupling settings on each channel		
Channel-to-Channel Crosstalk, typical	\geq 100:1 at 50 MHz, measured on one channel, with test signal applied between signal and common of the other channel, and with the same VOLTS/DIV and coupling settings on each channel		
Common to Chassis Capacitance, typical	65 pF		

Vertical				
Number of Channels	2			
Digitizers	8 bit resolution, separate digitizers for each channel sample simultaneously			
VOLTS/DIV Range	5 mV/div to 50 V/div at input BN0	C		
Polarity	Normal and Invert			
Position Range	±10 divisions	±10 divisions		
Analog Bandwidth	THS 710	THS 720		
at BNC, DC Coupled (at 5 mV/div_typical)	60 MHz at input BNC	100 MHz at input BNC		
		(90 MHz above 35° C)		
Peak Detect or Enve-	THS 710	THS 720		
lope Bandwidth, typical	50 MHz (1 µs/div or slower)	75 MHz (1 μs/div or slower)		
Analog Bandwidth Limit, typical	Selectable between 20 MHz or full			
Lower Frequency Limit, AC Coupled, typical	\leq 10 Hz at BNC, reduced by a factor of ten when using a 10X passive probe			
Rise Time at BNC,	THS 710	THS 720		
typical	5.8 ns	3.5 ns		
Peak Detect or Enve- lope Pulse Response, typical	Captures 50% or greater amplitude of pulses ≥ 8 ns wide (≥ 20 ns wide at 500 ns/div)			
DC Gain Accuracy	±2% for Sample or Average acquisition mode			
Position Accuracy	$\pm [0.4\% \times (\text{position} \times \text{volts/div}) + (0.1 \text{ div} \times \text{volts/div})]$			

Vertical		
DC Measurement	Measurement Type	Accuracy
Accuracy, Average Acquisition Mode	Average of ≥16 waveforms	\pm [2% × reading + (position × volts/div) + (0.1 div × volts/div)]
	Delta volts between any two averages of ≥16 waveforms acquired under same setup and ambient conditions	\pm [2% × reading + (0.05 div × volts/div)]
DC Measurement Accuracy, Sample Acq. Mode, typical	Acq. $\begin{vmatrix} \pm [2\% \times reading + (position \times volts/div) + (0.15 div \times v) \\ 0.6 mV \end{vmatrix}$	
Horizontal		
Sample Rate Range	THS 710	THS 720
Sample Rate Range	<i>THS 710</i> 5 S/s to 250 MS/s, in a 1.25, 2.5, 5 sequence	THS 720 5 S/s to 500 MS/s, in a 1.25, 2.5, 5 sequence
Sample Rate Range Record Length	THS 7105 S/s to 250 MS/s, in a 1.25, 2.5, 5 sequence2500 samples for each channel	<i>THS 720</i> 5 S/s to 500 MS/s, in a 1.25, 2.5, 5 sequence
Sample Rate Range Record Length SEC/DIV Range	THS 7105 S/s to 250 MS/s, in a 1.25, 2.5, 5 sequence2500 samples for each channelTHS 710	<i>THS 720</i> 5 S/s to 500 MS/s, in a 1.25, 2.5, 5 sequence <i>THS 720</i>
Sample Rate Range Record Length SEC/DIV Range (including MAG)	THS 7105 S/s to 250 MS/s, in a 1.25, 2.5, 5 sequence2500 samples for each channelTHS 71010 ns/div to 50 s/div	THS 720 5 S/s to 500 MS/s, in a 1.25, 2.5, 5 sequence THS 720 5 ns/div to 50 s/div
Sample Rate Range Record Length SEC/DIV Range (including MAG) Sample Rate and Delay Time Accuracy	THS 7105 S/s to 250 MS/s, in a 1.25, 2.5, 5 sequence2500 samples for each channelTHS 71010 ns/div to 50 s/div ± 200 ppm over any ≥ 1 ms time	THS 720 5 S/s to 500 MS/s, in a 1.25, 2.5, 5 sequence THS 720 5 ns/div to 50 s/div

Trigger			
Trigger Sensitivity, Edge Trigger Type	Coupling	Sensitivity	
	DC	0.35 div from DC to 50 MHz, increasing to 1 div at 100 MHz	
Trigger Sensitivity,	Coupling	Sensitivity	
Edge Trigger Type, typical	NOISE REJ	3.5 times the DC-coupled limits	
	HF REJ	1.5 times the DC-coupled limit from DC to 30 kHz, attenuates signals above 30 kHz	
	LF REJ	1.5 times the DC-coupled limits for frequencies above 1 kHz, attenuates signals below 1 kHz	
Trigger Level Range	±4 divisions from center of screen		
Trigger Level Accuracy, typical	$I_{,}\pm 0.2$ divisions, for signals having rise and fall times ≥ 20 ns		
SET LEVEL TO 50%, typical	Operates with input signals ≥50 Hz		
Width Range, Pulse Trigger Type, typical	99 ns to 1 s, with resolution of 33 ns or approximately 1% of setting (whichever is greater)		
Width Tolerance Range, Pulse Trigger Type, typical	5%, 10%, 15%, or 20%		

Oscilloscope Specifications	(Cont.)
-----------------------------	---------

Trigger		
Sensitivity, Video Trigger Type, typical	Composite video signal with negative sync pulse amplitude from 0.6 to 2.5 divisions	
Signal Formats and Field Rates, Video	Broadcast systems	Supports NTSC, PAL, and SECAM
Trigger Type	Interlaced	Field 1 or field 2
	Non-interlaced	Any field or any line
	Line Rates	15 kHz to 65 kHz, in five ranges
Holdoff Range	495 ns to 10 s	
Measurements		
Cursors	Voltage difference between cursors (Δ V) Time difference between cursors (Δ T) Reciprocal of Δ T in Hertz (1/ Δ T)	
Automated Measure- ments	Amplitude, Burst Width, Cycle Mean, Cycle RMS, Fall Time, Frequency, High, Low, Max, Mean, Min, Negative Duty Cycle, Negative Overshoot, Negative Width, Pk – Pk, Period, Positive Duty Cycle, Positive Overshoot, Positive Width, Rise Time, and RMS	

With P6113B Probe		
Analog Bandwidth, DC Coupled	THS 710	THS 720
	60 MHz	100 MHz
Probe Attenuation	10X	
Maximum Voltage Between Probe Tip and Reference Lead	Overvoltage Category	Maximum Voltage
	CAT II Environment	300 V _{RMS}
	CAT III Environment	150 V _{RMS}
	For steady-state sinusoidal waveforms, derate at 20 dB/decade above 100 kHz to 13 V_{pk} at 3 MHz and above. Also, refer to Overvoltage Category description on page A–15.	
Maximum Voltage Between Reference Lead and Earth Ground Using P6113B Probe	30 V _{RMS} , 42.4 V _{pk}	

Oscilloscope	Specifications	(Cont.)
--------------	-----------------------	---------

With P5102 Probe		
Analog Bandwidth, DC Coupled	THS 710	THS 720
	60 MHz	100 MHz
Probe Attenuation	10X	
Maximum Voltage Between Probe Tip and Reference Lead, DC Coupled	Overvoltage Category	Maximum Voltage
	CAT II Environment	1000 V _{RMS}
	CAT III Environment	600 V _{RMS}
Maximum Voltage Between Probe Tip and Reference Lead, AC Coupled	Overvoltage Category	Maximum Voltage
	CAT II Environment	±1000 V _{DC}
	CAT III Environment	±600 V _{DC}
Maximum Voltage	Overvoltage Category	Maximum Voltage
Between Reference Lead and Earth Ground	CAT II Environment	600 V _{RMS}
	CAT III Environment	300 V _{RMS}
Single Channel Com- mon Lead Feedthrough with P5102 Probe, typical	1000:1 from DC to 100 kHz, measured on either channel, with probe tip and reference lead connected together, and with test signal applied between tip/reference and earth ground A 3000 V/µs slew rate results in ≤0.5 division feedthrough, measured on either channel, with probe tip and reference lead connected together, and with test signal applied between tip/reference and earth ground	

DMM	Specifications
-----	-----------------------

General		
Resolution	$3\frac{3}{4}$ digit, 4000 count full scale reading except as noted	
Input Resistance, AC or DC Voltage	10 MΩ ±10%	
Input Capacitance, AC or DC Voltage, typical	≤100 pF	
Maximum Voltage	Overvoltage Category	Maximum Voltage
Between DMM and	CAT II Environment	600 V _{RMS}
	CAT III Environment	300 V _{RMS}
Maximum Voltage	Overvoltage Category	Maximum Voltage
Between DMM or COM Input and Earth Ground	CAT II Environment	600 V _{RMS}
	CAT III Environment	300 V _{RMS}
DC Voltage		
Ranges and Resolution	Range	Resolution
	400.0 mV	0.1 mV
	4.000 V	1 mV
	40.00 V	10 mV
	400.0 V	100 mV
	880 V	1 V

DMM Specifications (Cont.)

DC Voltage		
Accuracy	\pm (0.5% of reading + 5 counts)	
Normal Mode Rejec- tion, typical	Rejects AC signals by >60 dB at 50 Hz or 60 Hz (user selectable)	
Common Mode Rejec- tion, typical	Rejects AC signals by >100 dB at 50 Hz or 60 Hz (user selectable)	
AC Voltage		
Conversion Type	AC conversions are true RMS. The AC measurement is based on the AC and DC components of the signal as shown below:	
	AC Measurement = RMS(AC+DC) – DC	
Ranges and Resolution	Range	Resolution
	400.0 mV	0.1 mV
	4.000 V	1 mV
	40.00 V	10 mV
	400.0 V	100 mV
	640 V	1 V
Accuracy	Input Waveform	Maximum Error
	Sinusoidal waveforms with no DC component	\pm (2% of reading + 5 counts)
	Nonsinusoidal waveforms with crest factor up to 3 and no DC component	\pm (4% of reading + 5 counts)
Bandwidth, typical	≥5 kHz for all ranges	
Common Mode Rejec- tion, typical	Rejects AC signals by >60 dB at DC, 50 Hz, and 60 Hz	

Ω/Resistance		
Ranges and Resolution	Range	Resolution
	400.0 Ω	0.1 Ω
	4.000 kΩ	1Ω
	40.00 kΩ	10 Ω
	400.0 kΩ	100 Ω
	4.000 MΩ	1 kΩ
	40.00 MΩ	10 kΩ
Accuracy, typical	Range	Maximum Error
	All ranges except 40 M Ω	\pm (0.5% of reading + 2 counts)
	40 MΩ	\pm (2% of reading + 5 counts) for \leq 60% relative humidity
Bias Voltage for Full	Range	Full Scale Bias Voltage
Scale Resistance Mea- surement, typical	400.0 Ω	350 mV
	4.000 kΩ	200 mV
	40.00 kΩ	350 mV
	400.0 kΩ	350 mV
	4.000 MΩ	400 mV
	40.00 MΩ	1.10 V
Open Circuit Voltage, typical	Range	Open Circuit Voltage
	400.0 Ω	4.8 V
	All other ranges	≤1.2 V

DMM Specifications (Cont.)
DMM Specifications (Cont.)

Continuity Check		
Indication	An audible tone is generated when measured resistance is below 50 $\Omega_{\!\!\!,}$ typical	
Open Circuit Voltage, typical	4.8 V	
Test Current, typical	1 mA	
Diode Check		
Range	Zero to 2 V, measures forward voltage drop of semiconductor junction	
Voltage Accuracy, typical	±25%	
Open Circuit Voltage, typical	4.8 V	
Test Current, typical	1 mA	
Data Logger		
Horizontal Scale Range	24 hours/div to 30 s/div (4 minutes to 8 days, full scale)	

General Specifications

Display	
Display Type	4.7 in (120 mm) diagonal liquid crystal
Display Resolution	320 horizontal by 240 vertical pixels
Display Contrast	Adjustable, temperature compensated
Backlight Intensity, typical	35 cd/m ²

General Specifications	(Cont.)
-------------------------------	---------

RS-232 Interface			
Device Type	DTE		
RJ-45 Connector	Pin Number	Signal	
Pinout	1	DCD	
	2	DSR	
	3	DTR	
	4	GND	
	5	RXD	
	6	TXD	
	7	CTS	
	8	RTS	
Probe Compensator Output			
Output Voltage, typical	5.0 V into $\geq 1 M\Omega$ load		
Frequency, typical	1.2 kHz		
Power Source			
Battery	Replaceable Ni-Cd battery pack		
Battery Life, typical	Approximately two hours of continuous use from a full charge		
Low Battery Indication, typical	Low battery message first appears approximately ten minutes before the instrument powers off automatically		
Battery Saver	Standby Time-out and Backlight Time-out extend battery life. Time-out ranges from 1 minute to 15 minutes, or off.		
Battery Charging Time, typical	With TekScope instrument operating	20 hours	
	With TekScope instrument turned off	20 hours	
	In external charger	1.5 hours	

General Specifications (Cont.)

Power Source			
External Power	12 VDC nominal, center positive; Operates with input from 10 VDC to 15 VDC		
	The DC INPUT disconnects itself automatically if >15 VDC is applied. If this occurs, disconnect the overvoltage and then reconnect to a voltage in the proper range.		
Memory Retention, typical	All memory is retained indefinitely with battery removed and without external power applied.		
Fuse	This instrument has no user-replaceable fuses		
Environmental			
Temperature	Operating	–10° C to +50° C	
	Nonoperating	–20° C to +60° C	
Humidity	+40° C or below	≤95% relative humidity	
	+41° C to +50° C	≤75% relative humudity	
Altitude	Operating	2,000 m	
	Nonoperating	15,000 m	
Random Vibration	Operating	2.66 g _{RMS} from 5 Hz to 500 Hz, 10 minutes on each axis	
	Nonoperating	3.48 g _{RMS} from 5 Hz to 500 Hz, 10 minutes on each axis	
Drop Resistance, typical	Survives a 30 in drop onto concrete with only cosmetic damage		

General Speci	fications ((Cont.)
----------------------	--------------	--------

Mechanical		
Size	Height	8.53 in (217 mm)
	Width	6.95 in (177 mm)
	Depth	2.00 in (50.8 mm)
Weight	With battery installed	3.2 lbs (1.5 kg)
	With all standard accessories in soft carry case	7.5 lbs (3.4 kg)
	When packaged for domestic shipment	9.0 lbs (4.1 kg)
Certifications and Cor	npliances	
Certifications	Listed UL3111-1 and CAN/CSA-C22.2 No. 1010.1-92, complies with EN61010-1	
Overvoltage Category	Category	Examples
	CAT III	A typical CAT III environment is the power distribution system within a building or factory. These environments are some- what protected from lightning strikes, but susceptible to switching transients and other disturbances that may gener- ate high voltage impulses.
	CAT II	A typical CAT II environment is the 120/240 V distribution system within a lab or office. These environments are fairly well protected from external high voltage disturbances.

General Specifications (Cont.)

Certifications and Compliances			
EC Declaration of Conformity	Meets the intent of Directive 89/336/EEC for Electromagnetic Compatibility and Low-Voltage Directive 73/23/ECC for Product Safety. Compliance was demonstrated to the following specifications as listed in the official Journal of the European Communities:		
	EN 55	011 Class A:	Radiated and Conducted Emissions ¹³
	EN 50	081–1 Emissions: EN 60555–2	Power Harmonics
	EN 50	082–1 Immunity: IEC 801–2 IEC 801–3 IEC 801–4 IEC 801–5	Electrostatic Discharge RF Radiated ² Fast Transients Surge ³
	EN 61	010–1 Safety	
	1 T F	ektronix-supplied ferrite	bead required on instrument end of
	2 F n ≤	Performance criteria: \leq 5.0 div increase in peak-to-peak noise (Sample acquisition mode, full bandwidth); otherwise \leq 1.0 div increase in peak-to-peak noise	
	<mark>з</mark> д	Applies to instrument open dapter	erating from Tektronix-supplied AC
FCC Compliance	Emissions comply with FCC Code of Federal Regulations 47 CFR, Part 15, Subpart B, Class A		
Adjustment Interval			

The recommended adjustment interval is one year

Appendix B: Factory Setup

The table below lists the state of the TekScope instrument after you recall the Factory Setup.

Control	Changed by Factory Setup to
Acquire mode	Sample
Acquire stop after	HOLD button only
Acquire # of averages	16
Acquire # of envelopes	8
Channel selection	Channel 1 on, all others off
Cursor H Bar 1 position	-3.2 divisions from the center
Cursor H Bar 2 position	+3.2 divisions from the center
Cursor V Bar 1 position	-2 divisions from the center
Cursor V Bar 2 position	+2 divisions from the center
Cursor function	Off
Cursor time units	Seconds
Delayed time base time/div	50 µs/div
Delay time, delayed runs after main	200 ns
Data logger rate	30 s/div
Display format	YT
Display graticule type	Full
Display contrast	50%
Display style	Vectors
Display trigger "T"	On
Display accumulate time	500 ms

Control	Changed by Factory Setup to
DMM function	DC volts
DMM mode – autorange	Off
DMM mode – scope on/off	Off
Edge trigger coupling	DC
Edge trigger level	0.0 V
Edge trigger slope	Rising
Edge trigger source	Channel 1
Horizontal – main trigger position	50%
Horizontal – Mag	Off
Horizontal – time base	Main only
Main time base time/div	500 μs/div
Math waveform function	CH1 + CH2
Measure High-Low Setup	Histogram
Saved setups	No change
Saved waveforms	No change
Scope/DMM mode	Scope mode
Scope mode – autorange	Off
Scope mode – meter on/off	On
Trigger holdoff	Minimum (495 ns)
Trigger mode	Auto
Trigger type	Edge
Vertical bandwidth (all channels)	Full
Vertical coupling (all channels)	DC
Vertical position (all channels)	0 div
Vertical volts/div. (all channels)	100 mV/div

Appendix C: Accessories

Standard Accessories

P6113B 10X Passive Probes

The P6113B 10X passive probes have 100 MHz bandwidth and a CAT II voltage rating of 300 V _{RMS} . These probes are suitable for floating measurements up to 30 V _{RMS} .
--

Meter Lead Set

The standard pair of meter leads (012-1482-00) provides sharp tips for probing, two screw-on plunger tips for grabbing test points or small conductors, and two screw-on insulated alligator clips for grabbing terminals or larger conductors.

Battery Pack

AC Power Adapter

OF
OF

The AC power adapter allows operation from the AC power line and charges the internal battery. (North American 119-4812-00, Universal European 119-4813-00, United Kingdom 119-4922-00, Japan 119-4923-00)

RS-232 Cable and Adapter

	This RS-232 cable set includes a 2 m cable (012-1364-00) with RJ-45 connectors on each end. The set also includes a 9-pin adapter (103-0403-00) to connect to PCs.
--	--

Standard Accessories (Cont.)

Soft Case	
TEKSCOPE	The soft case (016-1399-00) protects the TekScope instrument when not in use. The soft case provides compartments for probes, one spare battery, AC power adaptor, and the User Manual.
Manual	

Optional Accessories

P5102 10X Passive Probes			
The P5102 10X passive probes have 100 MHz bandwidth a CAT II voltage rating of 1000 V _{RMS} . These probes are suita floating measurements up to 600 V _{RMS} .			
Deluxe Meter Lead Set			
	The deluxe pair of meter leads (012-1483-00) uses a sheathed banana-jack interface that is compatible with a variery of probing accessories. The deluxe set includes two sharp tips for probing, two plunger tips for grabbing component leads, one plunger tip for grabbing test points or small conductors, and one plunger tip with crocodile jaws for grabbing terminals or larger conductors. The cables have heat-resistant silicone insulation.		

Optional Accessories (Cont.)

THS7CHG Battery Charger



The battery charger recharges the battery pack in 1.5 hours. It can be powered from AC power or from the 12 V from an automobile cigarette lighter.

THS7BAT Extra Rechargeable Battery Pack



An extra rechargeable battery pack provides a high-capacity (4.8 V, $2.8 \text{ A}\cdot\text{hr}$), spare battery for extended portable operation.

Cigarette Lighter Adapter

$\mathbb{O}^{\mathbb{O}}$	
---------------------------	--

The cigarette lighter adapter (174-1734-00) allows you to operate the TekScope instrument or charge the internal battery from an automobile cigarette lighter.

HC 411 Thermal Printer



The HC 411 is a lightweight, portable, thermal printer that operates from AC or battery power and prints on 112 mm (4.4 in) wide paper. Additional paper is available; order part number 006-7580-00 for a package of five rolls.

Optional Accessories (Cont.)

THS7HCA Hard Carrying Case

The hard carying case provides ultimate protection for the TekScope instrument. This case has room for the TekScope instrument, voltage probes, meter leads, current probes, AC power adapter, battery charger, spare battery, and manuals.

A621 and A622 Current Probes



Two Tektronix current probes extend the TekScope instrument to handle current and power measurements. These clamp-on probes do not require disconnections to insert them into the circuit. Both current probes have BNC connectors for the oscilloscope channels and include safety banana-jack adapters for use with the DMM.



A621: AC only, 2,000 A max, 5 Hz to 50 kHz, selectable output at 1, 10, or 100 mV/A.

A622: AC or DC, 100 A max, DC to 100 kHz, selectable output at 10 or 100 mV/A.

A605 and A610 Current Probes



Optional Accessories (Cont.)

P6129B 1X-10X Switchable Passive Probes



The P6129B 1X-10X switchable passive probes have 10 MHz / 100 MHz bandwidth and are optimized for applications that require high sensitivity. Rated at 420 V(DC + peak AC), these probes are suitable for floating measurements up to 42 V_{pk} .

P6561AS SMD Probes



The P6561AS 10X passive probes have 100 MHz bandwidth and support low-voltage (<42 V_{pk}) logic probing. The probes connect to IC clip adapters or have minature adapters for probing surface-mount devices. These probes are suitable for floating measurements up to 42 V_{pk} .

P6408 Word Recognizer Trigger Probes



The P6408 is a 16-bit word recognizer probe for TTL logic operating at clock rates up to 20 MHz. This probe is used for trigger event recognition only, not logic signal display.

S60DWAV DocuWave Software



DocuWave is a DOS application that helps interface the TekScope instrument to your PC. You can upload and download waveforms and setups. You can export acquired data to spreadsheets for further analysis, to word processors to integrate with your documentation, or to printers and plotters to make hard copies.

Manuals

	_ \$ _	_	
	1		
	Ē		
	17	-++++-	

Programmer Manual (070-9245-XX) provides information about remote-control operation.

Service Manual (070-9246-XX) provides information about maintenence and module-level repair.

Appendix D: Performance Verification

This appendix contains performance verification procedures for the specifications marked with the \nvdash symbol. The following equipment, or a suitable equivalent, is required to complete these procedures.

Description	Minimum Requirements	Examples
DC Voltage Source	60 mV to 800 V, ±0.1% accuracy	Wavetek 9100 Universal Calibration System with
AC Voltage Source	300 mV to 640 V, ±0.5% accuracy at 500 Hz	Module (Option 250)
Leveled Sine Wave Generator	50 kHz and 100 MHz, ±3% amplitude accuracy	Fluke 5500A Multi-prod- uct Calibrator with Oscil- loscope Calibration
Time Mark Generator	10 ms period, ±50 ppm accuracy	Option (Option 5500A-SC)
Banana to Banana Cable (two required)	Shielded banana jacks on each end	Tektronix Deluxe Meter Lead Set (012-1483-XX)
50 Ω BNC Cable	BNC male to BNC male, \approx 36 in (1 m) long	Tektronix part number 012-0482-XX
50 Ω Feedthrough Termination	BNC male and female connectors	Tektronix part number 011-0099-XX
Dual Banana to BNC Adapter	Banana plugs to BNC female	Tektronix part number 103-0090-XX

Test Record

Serial Number	Procedure Performed by	Date

Test	Passed	Failed
Self Test		

Oscilloscope Tests		Low Limit	Test Result	High Limit
Channel 1 DC Measurement Accuracy	5 mV/div	34.05 mV		35.95 mV
	500 mV/div	3.405 V		3.595 V
	2 V/div	13.62 V		14.38 V
	10 V/div	68.1 V		71.9 V
Channel 2 DC Measurement Accuracy	5 mV/div	34.05 mV		35.95 mV
	500 mV/div	3.405 V		3.595 V
	2 V/div	13.62 V		14.38 V
	10 V/div	68.1 V		71.9 V
Channel 1 Bandwidth		425 mV		_
Channel 2 Bandwidth		425 mV		-
Sample Rate and Delay Time Accuracy		-4 divs		+4 divs
Channel 1 Edge Trigger Sensitivity		Stable trigger		-
Channel 2 Edge Trigger Sensitivity		Stable trigger		_

DMM Tests		Low Limit	Test Result	High Limit
DC Voltage Accuracy	400 mV range, 60 mV input	59.2 mV		60.8 mV
	400 mV range, 360 mV input	357.7 mV		362.3 mV
	4 V range	3.577 V		3.623 V
	40 V range	35.77 V		36.23 V
	400 V range	357.7 V		362.3 V
	880 V range	783 V		801 V
AC Voltage	400 mV range	352.3 mV		367.7 mV
Accuracy	4 V range, 600 mV input	0.583 V		0.617 V
	4 V range, 3.6 V input	3.523 V		3.677 V
	40 V range	35.23 V		36.77 V
	400 V range	352.3 V		367.7 V
	640 V range	559 V		593 V

Performance Verification Procedures

Before beginning these procedures, two conditions must first be met:

- The TekScope instrument must have been operating continuously for ten minutes within the operating temperature range specified.
- You must perform the Compensate Signal Path operation described on page 3–57. If the operating temperature changes by more than 5° C, you must perform the Compensate Signal Path operation again.

The time required to complete the entire procedure is approximately one hour.



WARNING. Some procedures use hazardous voltages. To prevent electrical shock, always set voltage source outputs to 0 V before making or changing any interconnections.

Self Test

This procedure uses internal routines to verify that the TekScope instrument functions and passes its internal self tests. No test equipment or hookups are required. Start the self test with the following setup:

SCOPE	UTILITY	System	Diag
		Loop	Once
		Execute	—
		OK Run Test	

A dialog box displays the result when the self test completes. Press the **CLEAR MENU** button to continue operation.



Check DC Measurement Accuracy

- 1. Set the DC voltage source output level to 0 V.
- 2. Set up the TekScope instrument using the following steps:

METER SCOPE				TOGGLE +
SCOPE	SAVE/ RECALL	Recall Saved Setup	Recall Factory Setup	_
		OK Recall Factory	—	
	CH 1	_		
	VERTICAL MENU	Probe Type	Voltage Probe	Set to 1X
	ACQUIRE	Mode	Average	Set to 16
	MEAS	Select Measrmnt	Mean*	_
		OK Select Measrmnt	—	

- * You may need to press Select Page to display this selection.
- **3.** Connect the TekScope instrument to the DC voltage source as shown below.



- **4.** For each VOLTS/DIV setting listed below, perform the following steps:
 - **a.** Set the DC voltage source output level to the positive voltage listed and then record the mean measurement as V_{pos} .
 - **b.** Reverse the polarity of the DC voltage source and then record the mean measurement as V_{neg} .
 - c. Calculate $V_{diff} = V_{pos} V_{neg}$ and then compare V_{diff} to the accuracy limits in the table.

VOLTS/DIV Setting	DC Voltage Source Output Levels	Accuracy Limits for V _{diff}
5 mV/div	+17.5 mV, –17.5 mV	34.05 mV to 35.95 mV
500 mV/div	+1.75 V, –1.75 V	3.405 V to 3.595 V
2 V/div	+7.00 V, -7.00 V	13.62 V to 14.38 V
10 V/div	+35.0 V, -35.0 V	68.1 V to 71.9 V

- 5. Set DC voltage source output level to 0 V.
- 6. To check channel 2, repeat step 2 substituting CH 2 for CH 1.
- 7. Press CH 1 and WAVEFORM OFF to remove the channel 1 waveform from the display.
- **8.** Repeat steps 3 through 5, substituting CH 2 for CH 1, to complete the check of channel 2.



Check Channel 1 Bandwidth

1. Set up the TekScope instrument using the following steps:

METER Scope				TOGGLE +
SCOPE	SAVE/ RECALL	Recall Saved Setup	Recall Factory Setup	_
		OK Recall Factory	—	
	ACQUIRE	Mode	Average	Set to 16
	TRIGGER	Coupling	Noise Reject	_
	MEAS	High-Low Setup	Min-Max	
		Select Measrmnt	Pk-Pk*	_
		OK Select Measrmnt	_	

- * You may need to press Select Page to display this selection.
- **2.** Connect the TekScope instrument to the leveled sine wave generator as shown below.



THS 710 & THS 720 User Manual

- 3. Set the TekScope instrument VOLTS/DIV to 100 mV/div.
- 4. Set the TekScope instrument SEC/DIV to 10 μs/div.
- 5. Set the leveled sine wave generator frequency to 50 kHz.
- 6. Set the leveled sine wave generator output level so the peak-topeak measurement is between **599 mV** and **601 mV**.
- **7.** Set the leveled sine wave generator frequency to **60 MHz** if you are checking a THS 710 or to **100 MHz** if you are checking a THS 720.
- 8. Set the TekScope instrument SEC/DIV to 10 ns/div.
- 9. Check that the peak-to-peak measurement is \geq 425 mV.
- **10.** Proceed to the next test to check the channel 2 bandwidth.



Check Channel 2 Bandwidth

1. First check the channel 1 bandwidth using the previous test. Then, perform these additional steps to check the channel 2 bandwidth:



- * You may need to press Select Page to display this selection.
- **2.** Connect the TekScope instrument to the leveled sine wave generator as shown below.



- 3. Set the TekScope instrument VOLTS/DIV to 100 mV/div.
- 4. Set the TekScope instrument SEC/DIV to 10 µs/div.
- 5. Set the leveled sine wave generator frequency to 50 kHz.
- 6. Set the leveled sine wave generator output level so the peak-topeak measurement is between **599 mV** and **601 mV**.
- **7.** Set the leveled sine wave generator frequency to **60 MHz** if you are checking a THS 710 or to **100 MHz** if you are checking a THS 720.
- 8. Set the TekScope instrument SEC/DIV to 10 ns/div.
- 9. Check that the peak-to-peak measurement is \geq 425 mV.



Check Sample Rate and Delay Time Accuracy

1. Set up the TekScope instrument using the following steps:

METER (SCOPE)				TOGGLE +
SCOPE	SAVE/ RECALL	Recall Saved Setup	Recall Factory Setup	—
		OK Recall Factory	—	
	VERTICAL MENU	Probe Type	Voltage Probe	Set to 1X

2. Connect the TekScope instrument to the time mark generator as shown below.



- 3. Set the time mark generator period to 10 ms.
- 4. Set the TekScope instrument VOLTS/DIV to 500 mV/div.
- 5. Set the TekScope instrument SEC/DIV to 2 ms/div.
- **6.** Use the vertical POSITION rocker to center the test signal on screen.
- 7. Press SET LEVEL TO 50%.
- 8. Change the TekScope instrument setup using the following steps:

METER METER SCOPE				TOGGLE +
SCOPE	HORIZON- TAL MENU	Time Base	Delayed Runs After Main	—
	CLEAR MENU	_	—	

9. Set the delayed time base to 500 ms/div.

10. Change the TekScope instrument setup using the following steps:

METER METER				TOGGLE +
SCOPE	Horizon- Tal	Time Base	Delayed Runs After Main	Set delay time to 10 ms

- 11. Set the delayed time base SEC/DIV to 500 ns/div.
- 12. Check that the rising edge of the marker crosses the center horizontal graticule line within ± 4 divisions of center graticule.

NOTE. One division of displacement from graticule center corresponds to a 50 ppm time base error.



Check Channel 1 Edge Trigger Sensitivity

1. Set up the TekScope instrument using the following steps:

METER SCOPE				TOGGLE +
SCOPE	SAVE/ RECALL	Recall Saved Setup	Recall Factory Setup	_
		OK Recall Factory	—	
	ACQUIRE	Mode	Average	Set to 16
	TRIGGER	Mode	Normal	_
	MEAS	High-Low Setup	Min-Max	
		Select Measrmnt	Ampl*	
		OK Select Measrmnt	_	

- * You may need to press Select Page to display this selection.
- **2.** Connect the TekScope instrument to the leveled sine wave generator as shown below.



- 3. Set the leveled sine wave generator frequency to 100 MHz.
- 4. Set the TekScope instrument VOLTS/DIV to 500 mV/div.
- 5. Set the TekScope instrument SEC/DIV to 10 ns/div.
- Set the leveled sine wave generator output level to approximately 50 mV_{p-p} so that the measured amplitude is approximately 500 mV. (The measured amplitude can fluctuate around 500 mV.)
- **7.** Press **SET LEVEL TO 50%**. Adjust **TRIGGER LEVEL** as necessary and then check that triggering is stable.
- 8. Change the TekScope instrument setup using the following steps:

METER (SCOPE)			
SCOPE	TRIGGER	Slope	\ (falling edge)

- **9.** Press **SET LEVEL TO 50%**. Adjust **TRIGGER LEVEL** as necessary and then check that triggering is stable.
- **10.** Change the TekScope instrument setup using the following steps:

SCOPE	TRIGGER	Slope	l (rising edge)

11. Proceed to the next test to check the channel 2 edge trigger sensitivity.



Check Channel 2 Edge Trigger Sensitivity

1. First check the channel 1 edge trigger sensitivity using the previous test. Then, perform these additional steps to check the channel 2 edge trigger sensitivity.



- * You may need to press Select Page to display this selection.
- **2.** Connect the TekScope instrument to the leveled sine wave generator as shown below.



- 3. Set the leveled sine wave generator frequency to 100 MHz.
- 4. Set the TekScope instrument VOLTS/DIV to 500 mV/div.
- 5. Set the TekScope instrument SEC/DIV to 10 ns/div.
- Set the leveled sine wave generator output level to approximately 50 mV_{p-p} so that the measured amplitude is approximately 500 mV. (The measured amplitude can fluctuate around 500 mV.)
- **7.** Press **SET LEVEL TO 50%**. Adjust **TRIGGER LEVEL** as necessary and then check that triggering is stable.
- 8. Change the TekScope instrument setup using the following steps:



9. Press **SET LEVEL TO 50%**. Adjust **TRIGGER LEVEL** as necessary and then check that triggering is stable.



Check DC Voltage Accuracy

- 1. Set the DC voltage source output level to 0 V.
- 2. Set up the TekScope instrument using the following steps:



3. Connect the TekScope instrument to the DC voltage source as shown below.



4. For each range, set the DC voltage source output to the level listed below, and then compare the meter reading to the accuracy limits.

Range	DC Voltage Source Output Level	Accuracy Limits
400 mV	60 mV	59.2 mV to 60.8 mV
400 mV	360 mV	357.7 mV to 362.3 mV
4 V	3.6 V	3.577 V to 3.623 V
40 V	36 V	35.77 V to 36.23 V
400 V	360 V	357.7 V to 362.3 V
880 V	792 V	783 V to 801 V

5. Set the DC voltage source output level to 0 V.



Check AC Voltage Accuracy

- 1. Set the AC voltage source output level to 0 V.
- 2. Set up the TekScope instrument using the following steps:



3. Connect the TekScope instrument to the AC voltage source as shown below.



- 4. Set the AC voltage source output frequency to 500 Hz.
- **5.** For each range, set the AC voltage source output to the level listed below, and then compare the meter reading to the accuracy limits.

Range	AC Voltage Source Output Level	Accuracy Limits
400 mV	360 mV	352.3 mV to 367.7 mV
4 V	600 mV	0.583 V to 0.617 V
4 V	3.6 V	3.523 V to 3.677 V
40 V	36 V	35.23 V to 36.77 V
400 V	360 V	352.3 V to 367.7 V
640 V	576 V	559 V to 593 V

6. Set the AC voltage source output level to 0 V.

This completes the performance verification procedure.

Appendix E: General Care and Cleaning

General Care

Protect the TekScope instrument from adverse weather conditions. The instrument is not waterproof.

Do not store or leave the instrument where the LCD display will be exposed to direct sunlight for long periods of time.



CAUTION. To avoid damage to the TekScope instrument, do not expose it to sprays, liquids, or solvents.

Cleaning

Inspect the TekScope instrument as often as operating conditions require. To clean the instrument exterior, perform the following steps:

- 1. Remove loose dust on the outside of the instrument with a lint-free cloth. Use care to avoid scratching the clear plastic display filter.
- 2. Use a soft cloth or paper towel dampened with water to clean the instrument. You can use a 75% isopropyl alcohol solution for more efficient cleaning.



CAUTION. To avoid damage to the surface of the TekScope instrument, do not use any abrasive or chemical cleaning agents.
Glossary and Index

Glossary

+/- Rocker

The general-purpose rocker button on the front panel that you can use to set parameters. The specific parameter assigned to the $^{+}/_{-}$ rocker depends on other selections.

AC Coupling

A mode that blocks the DC component of a signal but passes the dynamic (AC) component of the signal. Useful for observing an AC signal that is normally riding on a DC signal.

Acquisition

The process of sampling signals from input channels, digitizing the samples, processing the results into data points, and assembling the data points into a waveform record. The waveform record is stored in memory.

Active Cursor

The cursor that moves when you adjust the $^{+/-}$ rocker. The @ readout on the display shows the position of the active cursor.

Aliasing

A false representation of a signal due to insufficient sampling of high frequencies or fast transitions. A condition that occurs when an oscilloscope digitizes at an effective sampling rate that is too slow to reproduce the input signal. The waveform displayed on the oscilloscope may have a lower frequency than the actual input signal.

Attenuation

The degree the amplitude of a signal is reduced when it passes through an attenuating device such as a probe or attenuator (the ratio of the input measure to the output measure). For example, a 10X probe attenuates, or reduces, the input voltage of a signal by a factor of 10.

Auto Trigger Mode

A trigger mode that causes the oscilloscope to automatically acquire if it does not detect a triggerable event.

Autorange

A DMM feature that automatically sets the range to the optimum setting to measure an input signal. Also, an oscilloscope feature that automatically produces a stable waveform of usable size. In both cases, autorange continues to change the instrument settings to track additional signal changes.

Average Acquisition Mode

A mode in which the oscilloscope acquires and displays a waveform that is the averaged result of several acquisitions. This reduces the apparent noise. The oscilloscope acquires data as in the sample mode and then averages it according to a specified number of averages.

Backlight

The illumination behind the liquid-crystal display.

Bezel Buttons

The row of buttons below the display that selects items in the menus.

Common Lead

The meter lead attached to the reference voltage in a measurement application. Because of the isolated channels, the meter common lead and the scope reference leads do not have to be connected to the same reference voltage.

Continuity

A test to verify electrical conductivity from one point to another.

Cursors

Paired markers that you can use to make measurements between two waveform locations. The oscilloscope displays the values (expressed in volts or time) of the position of the active cursor and the distance between the two cursors.

DC Coupling

A mode that passes both AC and DC signal components to the circuit. Available for both the trigger system and the vertical system.

Digital Real Time Digitizing

A digitizing technique that samples the input signal with a sample frequency of four to five times the oscilloscope bandwidth. Combined with (sinx)/x interpolation, all frequency components of the input up to the bandwidth are accurately displayed.

Digitizing

The process of converting a continuous analog signal such as a waveform to a set of discrete numbers representing the amplitude of the signal at specific points in time.

Diode Test

A test to verify polarity and measure the forward voltage drop of a semiconductor junction.

Display

The word used to refer to the screen or the LCD (liquid crystal display).

Edge Trigger

Triggering that occurs when the oscilloscope detects the source passing through a specified voltage level in a specified direction (the trigger slope).

Envelope Acquisition Mode

A mode in which the oscilloscope acquires and displays a waveform that shows the variation extremes of several acquisitions.

Floating Measurements

Voltage measurements where the reference voltage is not earth ground. The two oscilloscope inputs and the DMM input are capable of taking independent floating measurements.

Ground (GND) Coupling

Coupling option that disconnects the input signal from the vertical system.

Hard Copy

An electronic copy of the display in a format useable by a printer or plotter.

Holdoff

A specified amount of time that must elapse after a trigger signal before the trigger circuit will accept another trigger signal. Holdoff helps ensure a stable display.

Horizontal Bar Cursors

The two horizontal bars that you position to measure the voltage parameters of a waveform. The oscilloscope displays the value of the active (moveable) cursor with respect to ground and the voltage value between the bars.

Isolated Channels

The architecture of the oscilloscope and DMM inputs that allow independent floating measurements. Each input can have a different reference voltage.

Menu

A set of labels shown in the display to identify the functions of the bezel buttons. The specific menu contents depend on the menu button you press.

Normal Trigger Mode

A mode where the oscilloscope does not acquire a waveform record unless a valid trigger event occurs. It waits for a valid trigger event before acquiring waveform data.

Paired Cursors

Two cross-shaped cursors that automatically track the vertical values of a waveform when you adjust their horizontal positions. The oscilloscope displays the voltage value and time value between the paired cursors.

Pixel

A visible point on the display. The display is 320 pixels wide by 240 pixels high.

Pop-Up Menu

A submenu of a menu. Pop-up menus temporarily occupy part of the waveform display area and present choices associated with the menu item selected. You can cycle through the options in a pop-up menu by repeatedly pressing the menu button underneath the pop-up.

Pretrigger

The specified portion of the waveform record that contains data acquired before the trigger event.

Pulse Trigger

Triggering on events that you can qualify by time. The oscilloscope triggers when an incoming pulse width meets time criteria you have defined.

Record Length

The specified number of samples in a waveform.

Reference Lead

The oscilloscope lead attached to the reference voltage in a measurement application. Because of the isolated channels, the meter common lead and the scope reference leads do not have to be connected to the same reference voltage.

Reference Waveform

A saved wavefrom selected for display. You can display two reference waveforms as Ref A and Ref B.

Rocker

A two-position button on the front panel used to control functions such as volts/division and trigger level.

Roll Mode

An acquisition mode useful at slow horizontal scale settings. Roll mode allows you to view the waveform as it is acquired point-by-point. The waveform appears to roll across the display.

RS-232

The serial communication port used to connect to a hard-copy device, computer, controller, or terminal.

Sample Acquisition Mode

A mode in which the oscilloscope creates a record point by saving the first sample during each acquisition interval. This is the default mode of the acquisition system.

Sample Interval

The time interval between successive samples in a time base. For real-time digitizers, the sample interval is the reciprocal of the sample rate.

Sampling

The process of capturing an analog input, such as a voltage, at a discrete point in time and holding it constant so that it can be quantized.

Selected Waveform

The waveform on which all measurements are performed and which is affected by vertical position and scale adjustments.

Signal Path Compensation (SPC)

The ability of the oscilloscope to minimize the electrical offsets in the vertical, horizontal, and trigger amplifiers caused by ambient temperature changes and component aging. You should run SPC when the ambient temperature varies more than 5° C from the last SPC or before performing critical measurements.

Standby (STBY)

The off-like state when the instrument in not in use. Some circuits are active even while the instrument is in the standby state.

Tek Secure

A feature that erases all waveform and setup memory locations (setup memories are replaced with the factory setup). Then it checks each location to verify erasure. This feature is useful where the oscilloscope is used to gather security-sensitive data.

Time Base

The set of parameters that let you define the time and horizontal axis attributes of a waveform record. The time base determines when and how long to acquire record points.

Vertical Bar Cursors

The two vertical bars you position to measure the time parameter of a waveform record. The oscilloscope displays the value of the active (moveable) cursor with respect to trigger and the time value between the bars.

Video Trigger

Triggering on the sync pulse of a composite video signal.

XY Format

A display format that compares the voltage level of two waveform records point by point. It is useful for studying phase relationships between two waveforms.

YT Format

The conventional oscilloscope display format. It shows the voltage of a waveform record (on the vertical axis) as it varies over time (on the horizontal axis).

Index

Α

accessories, C-1 ACQUIRE button, 3–3 acquisition modes, 3-3status, 3-32, 3-41 amplitude measurement, 3-26 applications detecting glitches, 2-24 displaying unknown signal, 2–14 measuring continuity, 2–17 measuring diode junctions, 2–17 measuring frequency, 2–18 measuring power, 2-28 measuring propagation delay, 2 - 20measuring resistance, 2–16 monitoring power quality, 2–34 testing switching transistor, 2–26 triggering on missing pulse, 2–22 triggering on motor RPM, 2–30 triggering on video field, 2-36 triggering on video line, 2–38 using scope and DMM, 2-32 automatic measurements, 3-24 **AUTORANGE** button, 2-4 turning off, 3-10 turning on, 3–8 average statistic, 3-30 averaging, 3-3

В

backlight, 3–56 bandwidth limit, 3–60 bar graph, 3–36 battery capacity, 1–6 charging, 1–6 external charger, C–3 low battery indicator, 1–6 replacing, 1–5 beep, 3–30, 3–56 burst width measurement, 3–26 buttons, using, 2–1, 2–4

С

cleaning, E–1 CLEAR MENU button, 2–5 connectors, 1–4 continuity, 2–17 contrast, display, 3–13 conventions, vi coupling trigger, 3–48 vertical, 3–60 current probe description, C–4 scale factor, 3–60 using, 2–28, 2–30, 3–15 CURSOR button, 3–11 cursors application example, 2–20 readouts, 3–12, 3–33 using, 3–11 cycle mean measurement, 3–26 cycle RMS measurement, 3–26

D

data logger application example, 2–34 display, 3–34 saving and recalling, 3–39 scroll speed, 3–23 delayed time base, 3–22 $\Delta 0$ statistic, 3–30 diagnostics, 3–58 diode test, 2–17 DISPLAY button, 3–13 display contents meter mode, 3–31 scope mode, 3–40 dots, 3–13

Ε

edge trigger, how to use, 3–48 envelope, 3–3 external power connector location, 1–4 using, 1–7

F

factory setup detailed description, B-1 recalling, 3-38 fall time measurement, 3-26 features automatic measurements, 2–18 cursors, 2-20 data logger, 2-34 dot accumulate, 2-36 isolated channels, 2-10, 2-26 overview, 1-1 pulse trigger, 2-22, 2-24, 2-30 using DMM with scope, 2-32video trigger, 2-36, 2-38 waveform math, 2–28 floating measurements application example, 2-26 overview, 2-10 frequency measurement, 3-26 front panel, 2-1, 2-4 functional check, 1–9

G

graticule, 3–13

Η

HARD COPY button, 3–17 high level measurement, 3–26 high voltage, warning, 2–12 HOLD button, 3–19 holdoff, 3–49 HORIZONTAL MENU button, 3–20

inputs, 1–4 intensity, display, 3–13 isolated channels application example, 2–26 description, 2–10

L

losing saved information, 1–5 low battery indicator, 1–6

Μ

MAG button, 3–21 manual conventions, vi math waveform allowed units. 3–61 application example, 2–28 functions, 3–61 max measurement, 3–27 max statistic, 3-30, 3-36 max-min statistic, 3-30 mean measurement, 3–26, 3–27 MEAS button, 3–24 measurement applications continuity, 2-17 data logger, 2–34 diode junction, 2-17floating, 2–10, 2–26 frequency, 2–18 motor current, 2–30 power, 2–28 propagation delay, 2–20 resistance, 2-16

measurement definitions meter mode, 3–30 scope mode, 3–26 measurement qualifier, 3–45 menus acquire, 3–3 cursor, 3-11 display, 3-13 horizontal, 3–20 how to use, 2-1measure, 3-24 meter, 3-31 save/recall, 3-37 trigger, 3-46 utility, 3–53 vertical, 3–59 meter deluxe lead set, C-2 lead set, C-1 noise reject, 3–63 range, 3-35 METER button, 3–31 min measurement, 3–27 min statistic, 3-30, 3-36

Ν

negative duty cycle measurement, 3–27 negative overshoot measurement, 3–27 negative width measurement, 3–27 nonvolatile memory saving information into, 3–38 when replacing battery, 1–5

0

ON/STBY button, 2–5 using, 1–5, 1–6 overrange indicator, 3–32

Ρ

P5102 compensating, 2-8 description, C-2 specifications, A-8 P6113B compensating, 2-8 description, C-1 specifications, A-7 peak detect, 3–3 peak-to-peak measurement, 3-27 performance verification, D-1 period measurement, 3-28 position horizontal, 3-21 of reference waveforms, 3-62 trigger, 3–22 vertical, 3-59 positive duty cycle measurement, 3 - 28positive overshoot measurement, 3 - 28positive width measurement, 3–28 power off timeout, 3-56 pretrigger, 3–22 printing, 3–17 probe compensator location, 1–4 using, 2–8

probe descriptions A605, A610, C-4 A621, A622, C-4 P5102, C-2 P6113B, C-1 P6129B, C-5 P6408, C-5 P6561AS, C-5 product description accessories, C-1 general, 1-1 meter, 1-3oscilloscope, 1-2 specifications, A-1 pulse trigger application examples, 2–22, 2-24, 2-30how to use, 3-50

R

readouts cursors, 3–12 DMM reading, 3-33, 3-42 measurements, 3-45 time base, 3–44 trigger, 3-44 waveform, 3-43 reference lead, 2–12 reference waveforms, 3-62 Rel Δ , 3–7 reset data logger, 3–34 statistics, 3-30 resistance, 2-16 rise time measurement, 3–28 RMS measurement, 3–28 roll mode, 3-21

RS-232 adapters, C–3 cable, C–3 hard copy, 3–17 port location, 1–4 setting up, 3–55 troubleshooting, 3–55 RUN/STOP button, 3–19

S

sample, 3–3 SAVE/RECALL button, 3–37 saved information, to prevent loss, 1 - 5SCOPE button, 3-40 SEC/DIV rocker meter mode, 3-23 scope mode, 3-21 selected waveform how to select, 3–59 indicator, 3-42 self test, 3–58 SET LEVEL TO 50% button, 2–5 setups factory setup description, B-1 saving and recalling, 3-37 signal path compensation, 3–57 single acquisition sequence, 3–6, 3 - 19specifications, A-1 statistics readout location, 3-33 resetting, 3–19 selecting, 3–29 status, 3-41 STBY. See ON/STBY

Т

T symbol, 3–14 Tek Secure, 3-54 tilt stand, 1–8 time base, 3-22time-out, 3-56 TOGGLE button, 3–11 TRIGGER MENU button, 3–46 trigger position, 3–22 trigger status, 3–41 triggering Auto mode, 3-48 edge trigger, 3-48 Normal mode, 3–48 on glitches, 2-25 on missing pulse, 2–22 on specific frequency, 2-30 on video field, 2–36 on video line, 2–38 pulse trigger, 3–50 readout, 3-44 setting coupling, 3-48 setting holdoff, 3-49 setting level, 3–46 setting pretrigger, 3–22 setting slope, 3–48 video trigger, 3-52

U

UTILITY button, 3–53

V

vectors, 3–13 VERTICAL MENU button, 3–59 video trigger application examples, 2–36, 2–38 how to use, 3–52 VOLTS/DIV rocker channels, 3–60 math waveform, 3–61 meter mode, 3–35

W

WAVEFORM OFF button, 3–59 waveforms, saving and recalling, 3–37

Χ

XY display, 3–14