

ABN 43 064 478 842

231 osborne avenue clayton south, vic 3169
 PO box 1548, clayton south, vic 3169
 t 03 9265 7400 f 03 9558 0875
 freecall 1800 680 680

www.tmgtestequipment.com.au

## Test & Measurement

- sales
- rentals
- calibration
- repair
- disposal

### **Complimentary Reference Material**


This PDF has been made available as a complimentary service for you to assist in evaluating this model for your testing requirements.

TMG offers a wide range of test equipment solutions, from renting short to long term, buying refurbished and purchasing new. Financing options, such as Financial Rental, and Leasing are also available on application.

TMG will assist if you are unsure whether this model will suit your requirements.

Call TMG if you need to organise repair and/or calibrate your unit.

If you click on the "Click-to-Call" logo below, you can all us for FREE!



Product Lifecycle Management System

#### Disclaimer:

All trademarks appearing within this PDF are trademarks of their respective owners.











- Modular, expandable system
- Surge voltage to 4.4 kV
- EFT/Burst to 4.8 kV/1 MHz
- PQT to 16 A/260 VAC & DC
- Easy to use 7" color touch screen
- TA (Test Assistance) provides fast standard test settings
- Parameters can be changed while test is running
- Wide range of optional test accessories

**Teseq's new NSG 3040 is an easy-to-use multifunction generator** that simulates electromagnetic interference effects for immunity testing in conformity with international, national and manufacturers' standards including the latest IEC/EN standards. The NSG 3040 system is designed to fulfill conducted EMC test requirements for CE mark testing, which generally include combination wave surge, Electrical Fast Transient (EFT) pulses and Power Quality Testing (PQT). Extensive expansion capabilities enable the system to be configured for a much broader range of applications.

**Featuring an innovative, modular design,** the NSG 3040 is a versatile system that can be configured for basic testing needs and expanded to meet the needs of sophisticated test laboratories. Teseq's well proven "Master-Slave" architecture enables individual pulse modules to be calibrated separately, with calibration data and correction factors stored on the slave controller. New modules can be easily installed with no need to return the entire system for calibration.

**Using state-of-the-art components,** the self-contained modules set new standards with respect to switching and phase accuracy and exceed the existing standards' requirements.

A 7" touch panel display with superb contrast and color makes controlling the NSG 3040 easy. For fast and efficient data entry, input devices include an integrated keyboard and a thumbwheel with additional keys for sensitivity adjustment. To achieve quick, reliable results in a development environment a standardized test can be initiated with just a few "clicks" using the integrated Test Assistance (TA) function.

**Convenient touch input buttons** make each parameter's value highly visible and allow the user to quickly select and modify all settings. A stylus is not necessary, and ramp functions can be programmed quickly and easily. Multi-step test procedures can be created and their sequence or parameter values can be changed easily.

**With expert mode users can make manual parameter changes** using the thumbwheel while a test is under way, providing an effective and fast method for identifying critical threshold values.

**An easily accessible SD memory card** allows firmware downloads to be performed quickly and tests to be saved. In the rare case that the storage space is not sufficient, the card can be replaced by a commercially available SD memory card and existing test files can be easily copied onto the larger SD card.

**The NSG 3040 has an Ethernet port for external PC control.** The Windows-based control software simplifies test programming and compilation of complex test sequences with various types of tests. Test reports can be generated during the test operation, allowing the operator to enter observations as the test progresses and increasing the efficiency of long-term tests.



#### The NSG 3040 performs tests according to the following specifications:

#### Combination wave pulse 1, 2/50 - 8/20 µs (Hybrid-Surge pulse)

Pulse conforms to IEC/EN 61000-4-5

| Parameter                      | Value   |
|--------------------------------|---|
| Pulse voltage (open circuit):  | ±200 V to 4.4 kV (in 1 V steps)                   |
| Pulse current (short circuit): | ±100 A to 2.2 kA                                  |
| Impedance:                     | 2/12 Ω  |
| Polarity:                      | positive / negative / alternate                   |
| Pulse repetition:              | 10 s, up to 600 s (in 1 s steps)                  |
| Test duration:                 | 1 to 9999 pulses, continuous                      |
| Phase synchronization:         | asynchronous, synchronous 0 to 359° (in 1° steps) |
| Coupling:                      | external / internal                               |

### Burst (EFT) 5/50 ns

Pulse conforms to IEC/EN 61000-4-4

| Parameter              | Value   |
|------------------------|---|
| Pulse amplitude:       | ±200 V to 4.8 kV (in 1 V steps) - open circuit      |
|                        | $\pm 100$ V to 2.4 kV (50 $\Omega$ matching system) |
| Burst frequency:       | 100 Hz to 1000 kHz                                  |
| Polarity:              | positive / negative / alternate                     |
| Repetition time:       | 1 ms to 4200 s (70 min)                             |
| Burst time:            | 1 μs to 1999 s, single pulse, continuous            |
| Test duration:         | 1 s to 1000 h                                       |
| Phase synchronization: | asynchronous, synchronous 0 to 359° (in 1° steps)   |
| Coupling:              | external / internal                                 |



#### Dips & drops

conforms to IEC/EN 61000-4-11, IEC/EN 61000-4-29

| Parameter                       | Value   |
|---------------------------------|---|
| Dips & drops:                   | From EUT voltage input to 0 V, 0%                 |
| Uvar with optional variac:      | depending on model (VAR 650x)                     |
| Uvar with step transformer:     | 0, 40, 70, 80% (INA 650x)                         |
| Peak inrush current capability: | 500 A (at 230 V)                                  |
| Switching times:                | 1 to 5 μs (100 Ω load)                            |
| Event time:                     | 20 µs to 1999 s, 1 to 99'999 cycles               |
| Test duration:                  | 1 s to 70'000 min, 1 to 99'999 events, continuous |
| Repetition time:                | 40 μs to 35 min, 1 to 99'999 cycles               |
| Phase synchronization:          | asynchronous, synchronous 0 to 359° (in 1° steps) |

### Variation test (with VAR 65xx only)

conforms to IEC/EN 61000-4-11

| Parameter                  | Value   |
|----------------------------|---|
| Uvar with optional variac: | 0 to 265 V (in 1 V steps), 0 to 115% (in 1% steps)              |
| Repetition time:           | 1 ms to 35 min, 1 to 99'999 cycles                              |
| Test duration:             | 1 ms to 5 s, 1 to 250 cycles (50 Hz);                           |
|                            | 1 to 300 cycles (60 Hz), abrupt                                 |
| Repetition time:           | 10 ms to 10 s, 1 to 250 cycles (50 Hz), 1 to 300 cycles (60 Hz) |
| Test duration:             | 1 s to 99'999 min, 1 to 99'999 events, continuous               |
| Phase synchronization:     | asynchronous, synchronous 0 to 359° (in 1° steps)               |

#### Pulsed magnetic field in conjunction with INA 753 and INA 701 or 702

conforms to IEC/EN 61000-4-9

| Parameter              | Value   |
|------------------------|---|
| Field:                 | 1 to 1200 A/m (in 1 A/m steps)                    |
| Polarity:              | positive / negative / alternate                   |
| Repetition time:       | 5 s to 10 min (in 1 s steps)                      |
| Impedance:             | 2 Ω   |
| Coil factor            | 0.01 to 50.00                                     |
| Test duration:         | 1 to 9'999 pulses, continuous                     |
| Phase synchronization: | asynchronous, synchronous 0 to 359° (in 1° steps) |



### Power magnetic field in conjunction with MFO 6501 / MFO 6502 and INA 70x conforms to IEC/EN 61000-4-8

| Field:         | 1 to max. 40 A/m (in 1 A/m steps) |
|----------------|-----------------------------------|
| Frequency:     | 50/60 Hz                          |
| Coil factor:   | 0.01 to 99.99                     |
| Test duration: | 1 to 9'999 pulses, continuous     |

#### Internal coupling network

| Parameter               | Value  |                                  |
|-------------------------|--|----------------------------------|
| Decoupling attenuation: | Remanent pulse 15% ma  | ax.                              |
| . 0                     | Mains side crosstalk 15% max.                                    |                                  |
| Mains decoupling:       | 1.5 mH 0% + 35%  |                                  |
| Connections:            | Back panel:  |                                  |
|                         | EUT supply: Harting con  | nector                           |
|                         | Additional ground conne  |                                  |
|                         | Instrument supply 230/1  | 115 VAC                          |
|                         | Front panel:   |                                  |
|                         | EUT connector IEC 320  |                                  |
|                         | HV coaxial   |                                  |
| EUT.                    | Connector surge high &   | low                              |
| EUT supply:             | 1-phase  | - (-)                            |
| EUT VAC:                |  | z (phase - neutral), 400 Hz max. |
| EUT VDC:                | 0 to 260 VDC   | /to man a vature a a catvalla d\ |
| EUT current             | 1 x 16 Arms continuous<br>1 x 25 Arms for 15 min                 | (temperature controlled)         |
| EFT (Burst)             | Standard coupling all lines to ref ground (GND) IFC/FN 61000-4-4 |                                  |
|                         | L, N, PE   | ⇒ ref GND                        |
|                         | Any lines and combinati  |                                  |
|                         |  | ⇒ ref GND                        |
|                         | N  | ⇒ ref GND                        |
|                         | PF   | ⇒ ref GND                        |
|                         | L, N   | ⇒ ref GND                        |
|                         | L, PE  | ⇒ ref GND                        |
|                         | N, PE  | ⇒ ref GND                        |
| PQT:                    | Dips & drops to phase L  |                                  |



#### Dimensions/weight

| Dimensions NSG 3040: | 449 (17.7") x 226 (8.9"; 5 HU) x 565 mm (22.2"), W x H x D |
|----------------------|--|
| Weight NSG 3040:     | approx. 25 kg (55 lbs)                                     |

#### **Options**

| Туре          | Description   |
|---------------|---|
| CDN 8014/8015 | Capacitive coupling clamp for burst                                 |
| CDN 163       | Burst coupling network 100 A per phase (coupling all to ref ground) |
| CDN 117/118   | Coupling networks for signal-/data lines (surge)                    |
| CAS 3025      | Burst/EFT verification set  |
| MD 200        | Voltage differential probe 7 kV                                     |
| MD 300        | Current probe 5 kA  |
| INA 165       | Conducted stand-off   |
| INA 166       | Brackets 5 HU for rack mounting                                     |

#### Accessories for IEC/EN 61000-4-11

| Туре     | Description                                     |
|----------|---|
| INA 6501 | Manual step transformer, 16 AAC, 0/40/70/80%    |
| INA 6502 | Automatic step transformer, 16 AAC, 0/40/70/80% |
| VAR 6501 | Automatic variable transformer, 7.5 A           |
| VAR 6502 | Automatic variable transformer, 2 x 16 A        |
| VAR 6503 | Manual variable transformer, 7.5 A              |

#### Accessories for IEC/EN 61000-4-8/-4-9

| Туре     | Description  |
|----------|--|
| MFO 6501 | Manual magnetic field option -4-8  |
| MFO 6502 | Automatic magnetic field option -4-8   |
| INA 701  | Magnetic field coil 1 x 1 m; with MFO max. 3.6 A/m -4-8;<br>Surge* max. 1200 A/m -4-9  |
| INA 702  | Magnetic field coil 1 x 1 m, with MFO max. 40 A/m -4-8;<br>Surge* max. 1200 A/m -4-9<br>*) Pulse shape adapter INA 753 needed to surge generator |
| INA 753  | Pulse shape adapter  |

